

In Vivo Photoacoustic Tomography of Chemicals: High-Resolution Functional and Molecular Optical Imaging at New Depths

Chulhong Kim, Christopher Favazza, and Lihong V. Wang*

Department of Biomedical Engineering, Washington University in St. Louis, Campus Box 1097, One Brookings Drive, St. Louis, Missouri 63130-4899

Received August 4, 2009

Contents

1. Introduction	2756	4.4. Fluorescence Proteins	2780
2. Photoacoustic Tomography Systems	2758	5. Summary	2780
2.1. Photoacoustic Computed Tomography (PACT)	2758	6. Acknowledgments	2781
2.1.1. A Single-Element Unfocused Ultrasound Transducer-Based PACT	2758	7. References	2781
2.1.2. A Ring Ultrasound Array-Based Real-Time PACT	2759		
2.1.3. A Linear Ultrasound Array-Based Real-Time PACT	2759		
2.2. Confocal Photoacoustic Microscopy (PAM)	2760		
2.2.1. A Single-Element Focused Ultrasound Transducer-Based Dark-Field Confocal PAM	2760		
2.2.2. Deep-Reflection Mode PA Macroscopy	2761		
2.2.3. Optical-Resolution PAM (OR-PAM)	2762		
3. Functional Photoacoustic Tomography Using Intrinsic Contrasts	2762		
3.1. Functional Imaging of Blood Vessels in Vivo Using PAT	2762		
3.2. Static and Dynamic SO ₂ Imaging in Rats' Skin	2763		
3.3. Imaging Brain Hemodynamic Changes in Small Animals	2764		
3.4. Imaging Tumor Angiogenesis and Tumor Hypoxia in Small Animals	2765		
3.5. Label-Free Functional Chronic Imaging of Wound Healing Process Using OR-PAM	2766		
4. Functional and Molecular PAT Using Exogenous Contrasts	2767		
4.1. Organic Dyes	2768		
4.1.1. Indocyanine Green (ICG)	2768		
4.1.2. IRDye-800 as Molecular Imaging Contrast Agents	2769		
4.1.3. Methylene Blue (MB) as a Contrast Agent on Sentinel Lymph Node (SLN) Mapping	2770		
4.2. Nanoparticles	2771		
4.2.1. Gold Nanocages	2771		
4.2.2. Gold Nanoshells	2772		
4.2.3. Gold Nanorods	2774		
4.2.4. Single-Walled Carbon Nanotubes	2777		
4.2.5. Colloidal Gold Nanobeacons	2778		
4.2.6. Nanowontons	2778		
4.2.7. Quantification of Optical Absorption Cross Sections of Gold Nanostructures Using Photoacoustic Sensing	2779		
4.3. Reporter Genes	2779		

1. Introduction

High-resolution volumetric optical imaging modalities, such as confocal microscopy, two-photon microscopy, and optical coherence tomography, have become increasingly important in the biomedical imaging field. However, due to strong light scattering, the penetration depths of these imaging modalities are limited to the optical transport mean free path in biological tissues, for example, ~1 mm in the skin. Photoacoustic tomography (PAT), an emerging hybrid imaging modality that can provide strong endogenous and exogenous optical absorption contrasts with high ultrasonic spatial resolution using the photoacoustic (PA) effect,¹ has overcome the fundamental depth limitation. The image resolution is scalable with the ultrasonic frequency. The imaging depth is limited to the reach of photons and up to a few centimeters deep in biological tissues. This Review will focus on the following aspects of PAT described in works published from 2003 to 2009: (1) multiscale PAT systems, (2) morphological and functional PAT using intrinsic contrasts (hemoglobin or melanin), and (3) functional and molecular PAT using exogenous contrast agents (organic dyes, nanoparticles, reporter genes, or fluorescence proteins).

Figure 1 illustrates the PA effect and its application to imaging. In general, a target is irradiated by a short-pulsed laser beam. A portion of the light is absorbed by the target's constituents and then partially converted into heat, which generates a pressure rise through thermo-elastic expansion. This initial pressure rise, determined by the local optical energy deposition and other thermal and mechanical properties, is propagated as an acoustic wave in the biological tissue and detected by ultrasonic transducers. PA images are derived from the arrival times of the ultrasonic waves. PAT image contrast is determined by optical absorption in the PA excitation phase, whereas its resolution is derived from ultrasonic detection in the PA emission phase. The combination of these intrinsic properties enables PA imaging to break through the fundamental limitations of pure optical imaging^{2–5} (i.e., shallow imaging depth for high resolution or low spatial resolution beyond the transport mean free path) and pure ultrasonic imaging (i.e., weak contrast for early cancer detection and speckle artifacts). By scaling the ultrasonic detection frequency, the spatial resolution and imaging depth can be tuned accordingly to the imaging application.

* To whom correspondence should be addressed. Tel.: (314) 935-6152. Fax: (314) 935-7448. E-mail: lhwang@seas.wustl.edu.



Dr. Chulhong Kim studied for his Ph.D. degree at Washington University in St. Louis, St. Louis, MO, under the supervision of Dr. Lihong V. Wang, Gene K. Beare Distinguished Professor. He is currently working at the Optical Imaging Laboratory as a postdoctoral research associate. His research interests are the development of novel biomedical imaging techniques including photoacoustic imaging, ultrasound-modulated optical tomography, fluorescence imaging, laser speckle contrast imaging, and ultrasonic imaging.



Dr. Christopher P. Favazza received his Ph.D. from Washington University in St. Louis, under the guidance of Dr. Ramki Kalyanaraman. He is currently an NIH postdoctoral research associate working in the Optical Imaging Laboratory at Washington University in St. Louis. His research interests include new clinical applications of biomedical imaging techniques (photoacoustic tomography, laser speckle contrast imaging, and ultrasound) in dermatology and microcirculation.

There are two major types of PAT systems. One is photoacoustic computed tomography (PACT), which uses reconstruction algorithms to generate an image.⁶ The second is photoacoustic microscopy (PAM) or macroscopy, which utilizes direct point by point detection and raster scanning over an object to render an image. PACT can be implemented using a single-element unfocused ultrasound transducer and circular scanning,⁶ or an ultrasound array system in circular,⁷ quarter-circular,⁸ or linear form. The imaging speed is dramatically improved in the latter approach and is ultimately limited to the pulse repetition rate of the excitation light source and the tissue damage threshold. By contrast, PAM does not require the use of an inverse algorithm to generate a three-dimensional image, but it requires raster scanning with a single-element focused transducer. An example of a PAM system is a dark-field confocal photoacoustic microscope.^{9,10} Based on recently reported experimental results, PAM imaging depth is scalable from 0.7 to 30 mm by varying the ultrasonic frequency from 75 to 3.5 MHz,^{10–12} whereas the spatial resolution is scalable from 5 to 150 μm .



Dr. Lihong Wang studied for his Ph.D. degree at Rice University, Houston, TX, under the tutelage of Drs. Robert Curl, Richard Smalley, and Frank Tittel. He currently holds the Gene K. Beare Distinguished Professorship in the Department of Biomedical Engineering at Washington University in St. Louis. He has authored and coauthored two books, including one of the first textbooks in the field of biomedical optics. He edited the first comprehensive book on biomedical photoacoustic tomography. He has published 206 peer-reviewed journal articles and delivered 228 keynote, plenary, and invited talks. He is a fellow of the American Institute for Medical and Biological Engineering, the Optical Society of America, the Institute of Electrical and Electronics Engineers, and the Society of Photo-Optical Instrumentation Engineers. He was appointed as the Editor-in-Chief of the Journal of Biomedical Optics. He serves as an equal cochair for the annual conference on Photons plus Ultrasound, the 2010 Gordon Conference on Lasers in Medicine and Biology, and the 2010 OSA Topical Meeting on Biomedical Optics. He serves as an equal cochair for the International Biomedical Optics Society. He has served as a study section chair or grant reviewer for the NIH and the NSF. He is currently a chartered member on an NIH study section. He serves as the founding chair of the scientific advisory board for a company commercializing his invention. His research on nonionizing biophotonic imaging has been funded with a cumulative budget of >\$26M (principal investigator for 21 research grants) by the NIH, NSF, and other funding agencies. He was a recipient of the NIH FIRST award and NSF CAREER award. His laboratory invented or discovered frequency-swept ultrasound-modulated optical tomography, dark-field confocal photoacoustic microscopy (PAM), optical-resolution PAM, photoacoustic Doppler sensing, photoacoustic reporter gene imaging, focused scanning microwave-induced thermoacoustic tomography, exact reconstruction algorithms for photoacoustic or thermoacoustic tomography, sonoluminescence tomography, Mueller-matrix optical coherence tomography, optical coherence computed tomography, and oblique-incidence reflectometry. In particular, PAM broke through the long-standing diffusion limit in penetration of conventional optical microscopy and reached superdepths for noninvasive biochemical, functional, and molecular imaging in living tissue at high resolution. His Monte Carlo model of photon transport in scattering media has been used worldwide.




Figure 1. Illustration of the PA mechanism. When a target absorbs light, PA waves are generated via thermo-elastic expansion. The PA waves propagate and are detected by an ultrasonic transducer.

Purely with endogenous contrasts such as hemoglobin or melanin, PAT is able to produce high-resolution images of vascular structures,^{9,12} solid tumors and angiogenesis,^{9,13} and internal organs^{11,14} in vivo. In addition to morphological information, this technique can provide functional informa-

tion such as oxygen saturation of hemoglobin and total hemoglobin concentration, using multiple wavelengths of light.^{6,9,15,16} Important to many functional measurements is the acquisition speed. With ultrasound array systems, PA imaging is sufficiently fast and capable of monitoring real-time heart beats in small animals.¹⁷ Exogenous contrast agents (organic dyes, nanoparticles, reporter genes, or fluorescence proteins) with biomarkers further enhance PA imaging, making it an increasingly valuable tool for molecular imaging.^{18,19}

2. Photoacoustic Tomography Systems

2.1. Photoacoustic Computed Tomography (PACT)

2.1.1. A Single-Element Unfocused Ultrasound Transducer-Based PACT

When the laser pulse width is shorter than the characteristic thermal diffusion time (referred to as thermal confinement), PA waves $p(\mathbf{r}_0, t)$ are related to light absorption $H(\mathbf{r}, t)$ as follows:²⁰

$$\frac{1}{\nu_s} \frac{\partial^2 p(\mathbf{r}_0, t)}{\partial t^2} - \nabla^2 p(\mathbf{r}_0, t) = \frac{\beta}{C_p} \frac{\partial H(\mathbf{r}, t)}{\partial t} \quad (1)$$

where ν_s is the acoustic velocity, C_p is the specific heat, and β is the volumetric thermal expansion coefficient. In addition, the light absorption $H(\mathbf{r}, t)$ is equal to $A_e(\mathbf{r})I(t)$, where $A_e(\mathbf{r})$ denotes the spatial distribution of optical energy deposition, $F(\mathbf{r})\mu_a(\mathbf{r})$, and $I(t)$ denotes the temporal laser pulse profile, where $F(\mathbf{r})$ denotes the local fluence and $\mu_a(\mathbf{r})$ denotes the local optical absorption coefficient.

If $I(t)$ is a Dirac delta function, a PA wave received at a detection point \mathbf{r}_0 (Figure 2) can be described as follows:²¹

$$p(\mathbf{r}_0, t) = \frac{\beta \nu_s^2}{4\pi C_p} \frac{\partial}{\partial t} \left[t \iint_{R=\nu_s t} A_e(\mathbf{r}) ds \right]$$

where ds is the surface element of a unit sphere and R is $|\mathbf{r} - \mathbf{r}_0|$.

By assuming the detection point \mathbf{r}_0 is far from the source point at \mathbf{r} as compared to the wavelength of PA wave, the spatial map of optical energy deposition $A_e(\mathbf{r})$ can be reconstructed as follows:²⁰

$$A_e(\mathbf{r}) = -\frac{2C_p}{\beta \nu_s^2} \int_{\Omega_0} t \frac{\partial p(\mathbf{r}_0, t = |\mathbf{r} - \mathbf{r}_0|/\nu_s)}{\partial t} d\Omega / \Omega_0 \quad (3)$$

where Ω_0 is the solid angle of the whole detection surface with respect to a given point source and $d\Omega$ is the solid angle




Figure 2. Circular scanning geometry in PACT using an unfocused ultrasound transducer.




Figure 3. A schematic of a PACT system for small-animal brain imaging. Reprinted with permission from ref 82. Copyright 2006 American Chemical Society.




Figure 4. Structural PACT image of a rat brain acquired in vivo. (A) Noninvasive PACT image of a rat brain cortex acquired with the skin and skull intact. C, cerebellum; H, cerebral hemispheres; O, olfactory bulbs; MCA, middle cerebral artery; CS, cruciate sulcus; FT, fissure transversa; V, blood vessels. (B) Open-skull photograph taken after PACT experiment. Reprinted with permission from ref 6. Copyright 2003 Nature.

of a detection element. If the fluence in the imaging plane is homogeneous at a given optical wavelength, the image of $A_e(\mathbf{r})$ is proportional to the image of $\mu_a(\mathbf{r})$; otherwise, fluence $F(\mathbf{r})$ needs to be considered for quantitative analysis of $\mu_a(\mathbf{r})$.

As an example, a schematic of a PACT system with a circularly scanning, single-element unfocused ultrasonic transducer used for small-animal brain imaging is shown in Figure 3.¹⁵ The system employed a tunable dye laser (ND6000, Continuum) pumped by an Nd:YAG laser (Brilliant B, Big Sky) to generate laser pulses with a pulse duration of 6.5 ns at 10 Hz. The laser beam was first expanded using a concave lens and homogenized using a ground glass. The incident energy density on skin surface was less than 3 mJ/cm², which is well below the ANSI standard of 20 mJ/cm².²² An unfocused ultrasonic transducer (e.g., XMS-310 or V383, Panametrics) was circularly scanned around the rat brain cortex in the horizontal (x - y) plane. By incrementally moving the ultrasonic transducer along the z -direction, volumetric information was obtained. The detected PA signals were first amplified and then digitized by an oscilloscope. Finally, 3D PA images were reconstructed and displayed in the computer. Because this system requires mechanical scanning of a single-element transducer, the scanning times required for both 2D and 3D images were approximately 15 min and several hours, respectively.

The PACT imaging was performed on an adult Sprague-Dawley rat (~350 g) using a 10 MHz ultrasonic transducer. The images were obtained noninvasively with the scalp and skull intact. Figure 4 shows the transdermal and transcranial images of a rat's brain acquired in vivo. The optical absorption is indicated with a gray scale, where the darker




Figure 5. A schematic of a 512-element, 360° PACT system. Reprinted with permission from ref 7. Copyright 2009 The Optical Society of America.




Figure 6. Time-resolved PA images of mouse brain vasculature with fast wavelength scanning. The laser wavelength was manually chirped from 710 to 900 nm over 30 s. Reprinted with permission from ref 7. Copyright 2009 The Optical Society of America.

areas, such as blood vessels, have relatively higher absorption. The PA image of the rat brain cortex (Figure 4A) matches well with the open-skull photograph (Figure 4B) obtained after PA imaging.

2.1.2. A Ring Ultrasound Array-Based Real-Time PACT^{7,8}

As mentioned above, single-element-based PACT systems can be relatively slow, and the example system required about 15 min to collect single 2D tomographic image. However, by employing a full-ring transducer array, the temporal resolution can be significantly improved. The system described below utilizes a 512 element full-ring transducer array and can collect one 2D tomographic image at the frame rate of 0.9 Hz.

A schematic of 512-element, 360° PACT system is shown in Figure 5. A Ti:Sapphire (Symphotics TII, LS-2134) laser pumped by a Q-switched Nd:YAG laser (Symphotics-TII, LS-2122) illuminated a small-animal brain with 8–12 ns pulses. The laser beam was first expanded and homogenized before striking the target. The custom-made transducer array (Imasonic) consisted of 512 elements in a curved aperture such that a complete circle (i.e., 360°) was formed with a radius of 25 mm. The center frequency of the transducers in this array system was 5 MHz with a nominal bandwidth greater than 80%. The system was also equipped with 512 dedicated low-noise and 20 dB gain preamplifiers. With 8:1 multiplexing, eight laser pulses were necessary to generate a complete 512-element capture. Data setup and transfer of results for a single acquisition could be performed at a frequency up to 8 Hz, leading to a maximum full-capture rate of about 1 frame/s.

The brain cortex of a mouse (~20 g) was noninvasively imaged to demonstrate the potential of dynamic spectroscopic PA imaging to monitor regional and temporal changes in oxygenation status throughout a small-animal brain. During

the 30 s imaging experiment, the laser wavelength of the laser was manually tuned from approximately 700 to 900 nm. Figure 6 shows the individual frames from the real-time sequence. The vascular structures of the brain cortex are clearly visible. The higher PA signals in blood vessels at 850 nm can be explained by both an increased penetration depth due to less optical scattering and higher absorption coefficients of the predominantly oxygenated hemoglobin at longer wavelengths. This experiment demonstrated the capability to perform wide-range, in vivo spectroscopic studies with subminute scan times.

2.1.3. A Linear Ultrasound Array-Based Real-Time PACT^{17,23,24}

Linear transducer arrays can also be used for PACT. A high-frequency (30 MHz) linear ultrasound transducer array was adapted to create a microscopic PACT system. This system performs real-time depth-resolvable 2D imaging (referred to as B-scan) at 50 Hz (faster than the 30 Hz video rate) with real-time beamforming and 3D imaging of 166 B-scan frames at 1 Hz post beamforming.

The schematic of microscopic PACT system is depicted in Figure 7. The tunable dye laser (Cobra, Sirah) pumped by a Nd:YLF laser (INNOSLAB, Edgewave) produced 7 ns laser pulses at the maximum repetition rate of 1 kHz without compromising the pulse energy. Dark-field confocal light delivery configuration was designed to provide a compact system with sufficient signal-to-noise ratio (SNR) and robust performance. The laser pulses were coupled into a multimode fiber and collimated at the exit of the fiber. The collimated beam was split into two beams by a 50:50 nonpolarizing beamsplitter. The two beams were reflected by mirrors toward two cylindrical lenses and coupled into a plastic slab. The optical fluence on the skin surface was ~2 mJ/cm² per pulse, well within the ANSI safety limit.²² A 30 MHz linear




Figure 7. Schematic of a microscopic PACT imaging system using a linear ultrasound array. Reprinted with permission from ref 23. Copyright 2008 Society of Photo-Optical Instrumentation Engineers.

ultrasound transducer array was fabricated by a 2–2-piezocomposite and was comprised of 48 elements ($82\ \mu\text{m} \times 2\ \text{mm}$) with $100\ \mu\text{m}$ spacing. The dimension of the element in the elevation direction was $2\ \text{mm}$, and the elements were focused in this direction with a fixed focal length of $8.2\ \text{mm}$. The axial, lateral, and elevational resolutions at $7\ \text{mm}$ depth from the transducer surface were estimated to be 25 , 70 , and $200\ \mu\text{m}$, respectively. The imaging depth in biological tissues was deeper than $3\ \text{mm}$. The detected PA signals were first amplified and then multiplexed to eight channels. Linear scanning was performed to achieve 3-D imaging.

To demonstrate this system's capability to image blood vessels *in vivo*, the upper dorsal region of a Sprague–Dawley rat ($\sim 95\ \text{g}$) was imaged. Figure 8 shows an *in vivo* PA image and an *ex vivo* transmission optical microscopic image of the subcutaneous vasculature in an imaged area. As shown in Figure 8, the vascular distributions in the two images matched well. The data acquisition for this image was completed within $1\ \text{s}$, and the MAP image was displayed within $3\text{--}4\ \text{s}$.

2.2. Confocal Photoacoustic Microscopy (PAM)

2.2.1. A Single-Element Focused Ultrasound Transducer-Based Dark-Field Confocal PAM^{9,10,25}

As opposed to PACT, dark-field confocal PAM utilizes a spherically focused ultrasound transducer. The axial resolution of this type of PAM is determined by the bandwidth of the transducer, and the transverse resolution is limited to the numerical aperture of the spherically focused acoustic lens. By measuring the time of arrivals of the PA signals, a one-dimensional depth resolved image can be obtained, referred to as an A-line. Raster scanning along the two transverse directions enables one to form three-dimensional images of optical absorption heterogeneities in biological tissues. Consequently, no additional reconstruction algorithm is required in PAM. The acquired volumetric images were expressed in one of two forms: a maximum amplitude projection (MAP), in which the maximum signal from each A-line is projected onto the corresponding XY plane, or a




Figure 8. (A) *In vivo* noninvasive PA image of subcutaneous blood vessels in a rat using microscopic PACT with a linear ultrasound array. The regions enclosed by major blood vessels are labeled using A–E. (B) Photograph taken from the dermal side of the excised skin with transmission illumination. Reprinted with permission from ref 23. Copyright 2008 Society of Photo-Optical Instrumentation Engineers.

true volumetric image processed by Volview software (Kitware Inc.).

A dark-field confocal PAM system is shown in Figure 9. The light was delivered from a tunable dye laser (ND6000, Continuum) pumped by a Q-switched Nd:YAG laser (LS-2137, LOTIS) with $6\ \text{ns}$ pulse duration and $10\ \text{Hz}$ pulse repetition rate. A multimode fiber was used for the light delivery. A dark-field illumination pattern was formed by the combination of a spherical conical lens and a weakly focusing optical condenser and was coaxially aligned with the ultrasound focus in water. This dark-field confocal configuration provides a greater penetration depth and higher SNR than the previously discussed confocal PACT system.²⁶ The light fluence on the skin was less than $6.4\ \text{mJ}/\text{cm}^2$, which is within the ANSI limit.²² Generated PA waves were detected by a single-element ultrasound transducer (20 or $50\ \text{MHz}$). The $50\ \text{MHz}$ ultrasonic frequency yielded 15 and $45\ \mu\text{m}$ axial and transverse resolutions, respectively. The maximum penetration depth at $50\ \text{MHz}$ was about $3\ \text{mm}$. As mentioned in the Introduction, the spatial resolution and imaging depth are highly scalable as a function of the ultrasonic frequency.

By taking advantage of the high optical absorption in hemoglobin and melanin, subcutaneous microvasculatures and melanoma in mice were imaged using the PAM system *in vivo*. Figure 10A shows a composite image of a melanoma and surrounding vascular structures. These dual pigments were imaged with two different wavelengths, $584\ \text{nm}$ for the surrounding microvasculature and $764\ \text{nm}$ for the melanoma. The bottom panel in Figure 10A shows the corresponding two-dimensional depth-resolved image (referred to as B-scan) along the dashed line in the top panel.




Figure 9. Experimental dark-field confocal PAM system. Reprinted with permission from ref 9. Copyright 2006 Nature.




Figure 10. In vivo noninvasive PA images of subcutaneous microvasculature, melanoma, and cortical vasculature in mice using dark-field PAM. (A) Top: A composite image of a melanoma acquired at 764 nm and surrounding microvasculature acquired at 584 nm using a 50 MHz ultrasonic detection frequency. Bottom: A B-scan image along the cut in the corresponding top figure. M, melanoma. (B) Top: A cortical vasculature image obtained at 570 nm using a 20 MHz ultrasonic detection frequency. Bottom: A B-scan image along the cut in the corresponding top figure. SS, sagittal sinus; MCA, middle cerebral artery; CS, coronal suture. Reprinted with permission from ref 9. Copyright 2006 Nature. Reprinted with permission from ref 27. Copyright 2009 American Institute of Physics.

The averaged image contrast of blood vessels at 584 nm as compared to background is 13 ± 0.89 , and that of the melanoma at 764 nm is 68 ± 5 . Figure 10B shows the mouse brain cortical microvasculature imaged using 20 MHz ultrasonic frequency and 570 nm optical wavelength.²⁷ Major vascular landmarks (sagittal sinus, middle cerebral artery, and coronal suture) are clearly identified. The corresponding bottom figure shows the B-scan image along the dashed line in the top figure.

2.2.2. Deep-Reflection Mode PA Macroscopy^{11,14}

To test the imaging capability in deep tissues and demonstrate the scalability of PAM, the deep-reflection mode PAM system was constructed and tested. The imaging penetration depth was increased more than 3 cm by lowering the ultrasonic frequencies to 3.5–10 MHz. For example, using a 5 MHz ultrasonic frequency, the spatial resolutions at 19 mm deep in biological tissue were 144 μm in the axial




Figure 11. In situ noninvasive PA images of internal organs in small and large animals using deep-reflection mode PAM. PA images of liver (A) and kidney (B) in a rat. PA images (C) of intestine in a rabbit, and the corresponding invasive photograph (D) taken after the imaging. AA, arcuate blood vessels; IB, interlobar blood vessels; IT, intestines; KN, kidney; LF, lumbodorsal fascia; LV, liver; RV, blood vessels around ribs; SC, blood vessels in the spinal cord; VC, vena cava inferior; BV, blood vessel; and IT, intestine. Reprinted with permission from ref 14. Copyright 2008 The American Association of Physicists in Medicine.

direction and 560 μm in the transverse direction. The principle of this system is the same as that of the previously described dark-field confocal PAM system,¹⁵ but the system has been scaled for deep imaging. Figure 11 shows the images of several internal organs such as a liver, a kidney, and an intestine from small and large animals imaged in situ. A Sprague–Dawley rat weighing ~ 200 g and a New Zealand white rabbit weighing ~ 3 kg were used for the small and large animal imaging, respectively. The liver and kidney of the rat are clearly identified in Figure 11A and B, respectively. The intestine of the rabbit is also clearly shown in Figure 11C and matches well with the photograph (Figure 11D) taken after the PA imaging.




Figure 12. Schematic of the OR-PAM. a, Condenser lens; b, pinhole; c, microscope objective; d, ultrasonic transducer; e, correcting lens; f, isosceles prism; g, acoustic lens; h, silicon oil. Reprinted with permission from ref 12. Copyright 2008 The Optical Society of America.

2.2.3. Optical-Resolution PAM (OR-PAM)¹²

Thus far, both axial and transverse resolutions in the above-mentioned PAM systems depend solely on ultrasonic parameters. To improve the ultrasonic resolution to the order of a micrometer, the penetration depth, which is limited by the frequency-dependent ultrasonic attenuation in tissues, becomes severely reduced. For example, to achieve a 5 μm transverse resolution, a 300 MHz of ultrasonic frequency is required. Consequently, the penetration depth is limited to $\sim 100 \mu\text{m}$ in tissue due to the strong acoustic attenuation. As an alternative approach, optical focusing can provide fine transverse resolution, while the axial resolution is still derived from the ultrasonic bandwidth. This technique is referred to as optical-resolution photoacoustic microscopy (OR-PAM). Although in this case the imaging depth is limited to the ballistic photon regime due to strong light scattering like other pure optical microscopy techniques,^{2–4} OR-PAM can still probe deeper than a PA system with the same transverse resolution provided by ultrasonic focusing. OR-PAM provides primarily optical absorption contrast, whereas conventional optical microscopy provides dominantly optical scattering or fluorescence contrast.

The experimental schematic of an OR-PAM system is shown in Figure 12. A dye laser pumped by a Nd:YLF laser was used as the PA excitation source. Laser pulses (pulse duration, 7 ns) passing through a 25 μm -diameter pinhole were spatially filtered, and then the pinhole was imaged to a nearly diffraction-limited focal spot with 3.7 μm in diameter through an objective lens (NA, 0.1; depth of focus, $\sim 40 \mu\text{m}$). About 100 nJ of the pulse energy after the objective lens was used. The ultrasound transducer (V2022 BC, Olympus NDT, 75 MHz) was focused through a plano-concave lens (radius of curvature, 5.2 mm; aperture, 6.4 mm; NA in water, 0.46). Therefore, both optical illumination and acoustic detection were coaxially aligned. A 3D image was acquired by raster scanning the surface of the target and combining the 1D depth scans received at each surface position. The transverse and axial resolutions of this system were 5 and 15 μm , respectively, and the maximum penetration depth was better than 700 μm .

Microvasculature in the ear of a nude mouse ($\sim 20 \text{ g}$) was imaged *in vivo* at an optical wavelength of 578 nm. Unlike that in ref 28, no optical clearing agent was applied to the skin surface. The microvasculature is clearly seen in two-dimensional MAP (Figure 13A) and volumetric OR-PAM images (Figure 11B), which agree well with the photograph (Figure 11C) taken with a transmission optical microscope. Small sizes of blood vessels are only observable in the OR-PAM image. The image contrast of blood vessels is nearly 20, as compared to the background. Some vessels labeled with CL show a diameter of $\sim 5 \mu\text{m}$, comparable to the diameter of a single capillary. These results demonstrate the capability of OR-PAM to image single capillaries *in vivo* without additional exogenous contrast agents.




Figure 13. OR-PAM images of microvasculature in a nude mouse ear *in vivo*. (A) 2D OR-PAM MAP image. (B) 3D visualization of the OR-PAM image. (C) Photograph taken with transmission optical microscopy. CL, capillary; and SG, sebaceous gland. Reprinted with permission from ref 12. Copyright 2008 The Optical Society of America.

3. Functional Photoacoustic Tomography Using Intrinsic Contrasts

3.1. Functional Imaging of Blood Vessels *In Vivo* Using PAT²⁹

Functional, biological images, such as the total hemoglobin concentration (HbT) and hemoglobin oxygen saturation (SO_2), are important in many medical and biomedical fields. These measurements are valuable for imaging brain activation,³⁰ studying tumor physiopathology,³¹ monitoring wound healing process,³² and studying gene expression.³³ Currently, the following other techniques can potentially estimate SO_2 :

(1) The first is near-infrared spectroscopy (NIRS).³⁴ NIRS measures reflected or transmitted diffusive light from biological tissues at various optical wavelengths and estimates the SO_2 comparing optical–spectral differences between deoxy-hemoglobin (HbR) and oxy-hemoglobin (HbO_2). However, it is difficult to collect localized measurements due to strong light scattering in biological tissues.

(2) The second is blood-oxygen-level-dependent contrast magnetic resonance imaging (BOLD MRI).³⁵ Despite the high spatial resolution of BOLD MRI, the contrast is only sensitive to HbR; consequently, measuring oxygenation and blood flow changes in vessels with this technique is challenging.

(3) The third is electron paramagnetic resonance imaging (EPRI).³⁶ EPRI requires the injection of exogenous free-radical contrast agents.




Figure 14. Spectra of two types of hemoglobin: oxy-hemoglobin (HbO₂) and deoxy-hemoglobin (HbR).

(4) Fourth is positron emission tomography (PET) and single photon emission tomography (SPET).^{37,38} Both techniques require the injection of radioactive isotopes and suffer from poor spatial resolution.

PA measurements of SO₂ use the same principle as NIRS, where hemoglobin is considered the dominant optical absorption chromophore at each wavelength (λ_i). The two forms of hemoglobin, HbR and HbO₂, have different molar extinction spectra (Figure 14), where the molar extinction coefficient is defined as the extinction coefficient of a substance at a specific unit of concentration.

Thus, as a first-order approximation, the blood absorption coefficient $\mu_a(\lambda_i)$ can be calculated as follows:

$$\mu_a(\lambda_1) = \varepsilon_{\text{ox}}(\lambda_1)C_{\text{ox}} + \varepsilon_{\text{de}}(\lambda_1)C_{\text{de}} \quad (4)$$

$$\mu_a(\lambda_2) = \varepsilon_{\text{ox}}(\lambda_2)C_{\text{ox}} + \varepsilon_{\text{de}}(\lambda_2)C_{\text{de}} \quad (5)$$

where μ_a (cm⁻¹) is the absorption coefficient; λ_1 and λ_2 are the two PA imaging wavelengths; ε_{ox} and ε_{de} are the known molar extinction coefficients (cm⁻¹ mM⁻¹) of the HbO₂ and HbR, respectively; and C_{ox} and C_{de} are the concentrations (mM) of the HbO₂ and HbR. The SO₂ and HbT are defined as follows:

$$\text{SO}_2 = \frac{C_{\text{ox}}}{(C_{\text{ox}} + C_{\text{de}})} \quad (6)$$

$$C_{\text{HbT}} = C_{\text{ox}} + C_{\text{de}} \quad (7)$$

Because the localized PA amplitudes are proportional to the local optical energy deposition, one can replace μ_a with the measured PA amplitudes. In addition, to minimize wavelength-dependent fluence fluctuation, close wavelengths (i.e., 566 and 570 nm) are usually selected. Many groups work on the correction of wavelength-dependent fluence fluctuation as well. It is shown that even a simple difference between two images acquired using two appropriately different wavelengths is correlated with the oxygenation of hemoglobin. More than two optical wavelengths, however, are often used to improve accuracy.

3.2. Static and Dynamic SO₂ Imaging in Rats' Skin^{9,16}

The 50 MHz dark-field confocal PAM (see section 2.2.1) was employed for both static and dynamic SO₂ imaging in Sprague–Dawley rats (~200 g) in vivo. Four optical wavelengths (578, 584, 590, and 596 nm) were applied sequentially to get multicolored PA images. Figure 15A shows the morphological MAP image of blood vessels taken at 584 nm (isosbestic wavelength), which reflects the total




Figure 15. Static and dynamic SO₂ imaging in rats' skin. (A) Structural MAP image taken at 584 nm. (B) Static SO₂ mapping. Red, arterial blood; blue, venous blood. (C) Structural MAP image taken at 584 nm. (D) Static SO₂ mapping under normoxia. Red, arterial blood; blue, venous blood. (E) Changes in SO₂ from normoxia to hypoxia. (F) Changes in SO₂ from normoxia to hyperoxia. (G) Quantification of SO₂ values in arteries and veins under three physiologic conditions. Reprinted with permission from ref 9. Copyright 2006 Nature. Reprinted with permission from ref 16. Copyright 2007 American Institute of Physics.

hemoglobin concentration. Vessel-by-vessel mapping of SO₂ using eq 6 is shown in Figure 15B, estimated from the four images. Red color represents arterial blood (SO₂ = 0.97 ± 0.02), and blue color stands for venous blood (SO₂ = 0.77 ± 0.04), and these values agree well with values previously reported in the literature.³⁹ With a high-resolution PAM system, vessel-by-vessel SO₂ mapping is possible, while the other techniques show only the volume-averaged SO₂.

For dynamic SO₂ mapping, three physiological conditions (hyperoxia, normoxia, and hypoxia) were induced in a rat by changing the oxygen concentration of the inhaled gases (pure oxygen, normal air, and carbogen comprising 5% O₂, 5% CO₂, and 90% N₂). The other imaging protocols remained the same as in the static imaging experiment. Figure 15C shows the morphological MAP image of blood vessels taken at 584 nm under hyperoxia. Figure 15D shows the SO₂ imaging under normoxia. Figure 15E shows the differential SO₂ image between normoxia and hypoxia (hypoxia – normoxia). In addition, the SO₂ changes between normoxia and hyperoxia (hyperoxia – normoxia) are shown in Figure 15F. The SO₂ values in the arteries and veins under three physiological conditions are quantified in Figure 15G. The changes in SO₂ between the three conditions agreed well with the pulse oximeter readings and previously reported results.⁴⁰

3.3. Imaging Brain Hemodynamic Changes in Small Animals^{6,15,41}

First, a single-element unfocused ultrasound transducer-based PACT (see section 2.1.1) was employed to assess the cerebral blood volume and SO_2 in brains of Sprague–Dawley rats (~ 80 g) *in vivo*. Three physiological states (hyperoxia, normoxia, and hypoxia) were induced in a rat by changing oxygen concentration in the inhaled gases as explained in section 3.2. Two PA images were acquired at two optical wavelengths (584 and 600 nm) for each physiological state. Figure 16A and B shows the morphological images of the brain cortex under normoxia taken at 584 and 600 nm, respectively, which match well with the photograph (Figure 16I) taken after the imaging experiments. PA signals in Figure 16A are stronger than those in Figure 16B because the optical absorption coefficients of both types of hemoglobin are higher at 584 nm than at 600 nm (Figure 14). For the HbT and SO_2 mapping using eqs 6 and 7, the cortical venous regions were segmented from the background. Figure 16C and D shows the mapping of SO_2 and HbT under normoxia, respectively. The SO_2 image indicates the absolute value, whereas the HbT image has an arbitrary unit. Figure 16E shows the differential $\langle \text{SO}_2 \rangle$ (the averaged SO_2 level) image between normoxia and hyperoxia (hyperoxia – normoxia). Additionally, the $\langle \text{SO}_2 \rangle$ changes between normoxia and hypoxia (hypoxia – normoxia) are shown in Figure 16F. Under hyperoxia, $\langle \text{SO}_2 \rangle$ in the cortical venous vessels was higher than that under normoxia, and as expected $\langle \text{SO}_2 \rangle$ under normoxia was higher than that under hypoxia. The quantified $\langle \text{SO}_2 \rangle$ values were $\sim 80\%$, $\sim 70\%$, and $\sim 57\%$ under hyperoxia, normoxia, and hypoxia, respectively. Simultaneously, changes in the averaged HbT value, $\langle \text{HbT} \rangle$, were estimated during the changes of physiological states. Change in $\langle \text{HbT} \rangle$, denoted by $\langle \Delta \text{HbT} \rangle / \langle \text{HbT} \rangle$, from normoxia to hyperoxia was $\sim 4\%$ decrease (Figure 16G), whereas from normoxia to hyperoxia was $\sim 12\%$ increase (Figure 16H). All segmented functional images in Figure 16E–H were overlaid on the structural image in Figure 16B. These results agree well with the previously reported results.⁴⁰

Using the PACT system, functional brain activation in rats' brain cortexes was studied in response to left and right whisker stimulations in Sprague–Dawley rats (~ 350 g). PA images acquired without whisker stimulation were subtracted from images acquired with whisker stimulation to map the brain activation. Figure 17A and B shows functional cerebral hemodynamic changes in response to left and right whisker stimulation, respectively. The activated regions in PA images were marked on the open-skull photograph with two boxes taken after the PA imaging experiments (Figure 17C). This region was histologically analyzed. The vibrotome sections of cortical layer IV, whisker-barrel cortex in Figure 17D, colocalized well with the activated regions in PA images. Increases in PA signals in those areas could be explained as increase in blood volume and/or flow due to the whisker stimulation.

The 20 MHz dark-field confocal PAM (see section 2.2.1) was also employed to image brain activity via the hemodynamic response in Swiss Webster mice (~ 25 g) *in vivo*. The real-time oxygenation dynamics of the selected individual cortical vessels were measured simultaneously in response to controlled perturbations of inhaled oxygen levels. Figure 18B shows the morphological PA image of the brain cortex in the marked region in Figure 18A at 570 nm (isosbestic wavelength). After several major somatosensory cortex (SC)




Figure 16. *In vivo* noninvasive PA images of rat brain hemodynamic changes including total hemoglobin concentration (HbT) and oxygen saturation of hemoglobin (SO_2). (A and B) Structural brain images under normoxia at 584 and 600 nm laser light, respectively. (C and D) Images of SO_2 and HbT in the cortical venous area under normoxia. (E) The differential $\langle \text{SO}_2 \rangle$ (the averaged SO_2 level) image between normoxia and hyperoxia (hyperoxia – normoxia). (F) The differential $\langle \text{SO}_2 \rangle$ image between normoxia and hypoxia (hypoxia – normoxia). (G) The fractional change in HbT ($\langle \Delta \text{HbT} \rangle / \langle \text{HbT} \rangle$) between hyperoxia and normoxia. (H) The fractional change in HbT between hypoxia and normoxia. (I) Open-skull anatomical photograph taken after the PA imaging experiments. Reprinted with permission from ref 15. Copyright 2006 Society of Photo-Optical Instrumentation Engineers.




Figure 17. Functional PA images of cerebral hemodynamic changes in response to whisker stimulations. (A and B) Cerebral functional PA images in response to left and right whisker stimulation, respectively. The map of increased PA signals (color) was superimposed on the gray scale structural PA image. (C) Open-skull photograph taken after the PA imaging experiments. B, bregma; L, lambda; M, midline; A, activated regions. (D) Histology of normal lamina IV cortical barrels (region A). Reprinted with permission from ref 6. Copyright 2003 Nature.




Figure 18. In vivo structural PA images of the mouse cortex vasculature. (A) Photograph of the scanned region. (B) Noninvasive PA image of the mouse cortex vasculature. SS, sagittal sinus; CS, coronal suture. (C) B-scan image obtained from scanning along the white dotted line in (B). Reprinted with permission from ref 41. Copyright 2009 Society of Photo-Optical Instrumentation Engineers.

vessels were identified in Figure 18B, a B-scan (Figure 18C) was obtained by raster scanning along the dashed line in Figure 18B. The temporal resolution of one B-scan (Figure 18C) is less than 1 s.

Two optical wavelengths (561 nm, deoxyhemoglobin dominant wavelength; 570 nm, an oxygen-insensitive absorption wavelength) were applied sequentially to get multicolored PA images. The acquisition time for a pair of dual wavelength scans (10 s) was limited by the wavelength tuning time of the dye laser. Throughout image acquisition, hyperoxic (pure oxygen: 100% O₂) and hypoxic (carbogen: 5% O₂, 5% CO₂, and 90% N₂) states were induced by alternating the inhaled oxygen levels every 6 min, and a pulse oximeter was used to verify each change. To estimate dynamic oxygenation changes, the integration of the full width at half-maximum (fwhm) of the image of each vessel over the time period was calculated to minimize the laser fluctuation effects. A hyperoxic state (100% O₂) was used as a baseline, and the animal was exposed to three 5% O₂ steps with a return to baseline between successive hypoxic states. Figure 19A shows the dynamic oxygenation profile for each of the five vessels, and a close-up of the profile is shown in Figure 19C to emphasize the transition phases between two physiological states. The baseline (isosbestic) measurement is shown in Figure 19B. The initial step from hyperoxia to hypoxia occurred at time $t = 6.1$ min. The forward response time, characterized by the rising time from




Figure 19. Dynamic oxygenation profiles of five vessels. Profiles acquired through a hypoxic challenge and shown in changes (%) of (A) ratiometric PA signals and (B) isosbestic PA signals. (C) Close-up of the forward rising part of the dynamic vessel responses shown in (A). Reprinted with permission from ref 41. Copyright 2009 Society of Photo-Optical Instrumentation Engineers.

10% to 90% of maximum, was 63 ± 6 s. The forward response time among the five vessels did not vary significantly ($\alpha = 0.1$). Although the maximum values under hypoxia significantly fluctuated among the vessels, no significant correlation between maximum response and vessel diameter was observed. At $t = 12.2$ min, the hypoxic challenge ended via a step exposure to 100% O₂. The reverse response time, defined as the falling time from 90% to 10% of the maximum, was 16 ± 2 s.

3.4. Imaging Tumor Angiogenesis and Tumor Hypoxia in Small Animals^{13,42}

Tumor angiogenesis^{43,44} describes the formation of new blood vessels around and within the tumor and is a critical step in tumor growth. Angiogenesis increases the blood and nutrient supply to the tumor, and consequently expedites tumor growth. During the normal angiogenesis process, the balance of pro- and antiangiogenic signals is tightly regulated, so new blood vessels become quickly mature and stable. However, during the tumor angiogenesis, this balance is not maintained, and new vessels are constantly formed. The structures of these blood vessels are different from normal vessels. They are typically abnormally shaped, dilated, tortuous, and can have blind ends. In addition, these vessels are leaky and hemorrhagic from overproduction of vascular endothelial growth factor (VEGF). Because of these differences, tumor angiogenesis is an important parameter for tumor diagnosis and can be detected with PAT. A single-element unfocused ultrasound transducer (3.5 or 20 MHz)-based PACT (see section 2.1.1) was employed to image tumor angiogenesis in tumor-bearing brains of Sprague–Dawley rats (~100–200 g) in vivo. The mammary adenocarcinoma cell line (*Br7–C5*)⁴⁵ was transcranially injected (orthotopic inoculation, 105 monodispersed cells) into the brain cortex. One month after the intracerebral inoculation, a tumor nodule (0.5 cm in diameter) developed on the brain cortex. Figure 20A and B shows two in vivo PA images of brain tumors, acquired noninvasively with 3.5 and 20MHz ultrasonic transducers, respectively. Two corresponding close-up images of the tumors are shown in Figure 20 as well. The images were compared to the open-skull photograph of the rat brain taken after the PA imaging in Figure 20C. Irregular blood




Figure 20. In vivo PA images of tumor angiogenesis and the corresponding close-up images in a rat brain acquired with (A) a 3.5 MHz ultrasonic transducer and (B) a 20 MHz ultrasonic transducer. (C) Photograph of the rat brain LH, left cerebral hemisphere; V, blood vessels; E, eyes; T, tumor. Reprinted with permission from ref 13. Copyright 2005 The Optical Society of America.

vessels with distorted shapes are easily recognized within the tumor shown in the two images. Furthermore, hemorrhagic infiltration of the parenchyma due to leaky blood vessels containing interendothelial junctions, an incomplete or absent basement membrane, and large numbers of transendothelial channels can clearly be observed in the region of the tumor.⁴⁶ In the close-up image acquired with the 20 MHz transducer, the vessels show better delineated boundaries, as expected.

Tumor hypoxia^{47,48} is a situation in which tumor cells lack oxygen. Some regions of the tumor have lower oxygen concentration than healthy tissue. Tumor hypoxia is primarily caused by unbalanced oxygen consumption as compared to oxygen supply due to the abnormal tumor vasculature. As described above, tumor vasculature is extremely abnormal and is marked by random vascular dilations and networks, elongated and tortuous structures, and lack of flow regulation. These effects cause severe deficiencies in the perfusion of oxygen and nutrients, leading to ischemic hypoxia. In addition, those cells in the hypoxic zone are quiescent and prevent not only oxygen but also anticancer agents from penetrating. Tumor hypoxia is a distinctive feature, important for tumor diagnosis, and can be imaged with PAT. A single-element unfocused ultrasound transducer (10 MHz)-based PACT (see section 2.1.1) was used to image tumor hypoxia in tumor-bearing brains of nude mice (~20 g) in vivo. Human U87 glioblastoma tumor cells were inoculated stereotactically into mice. One $\times 10^6$ cells of 7 μL were inoculated intracranially into the caudate nucleus at a 3 mm depth from the skin surface.⁴⁹ Four optical wavelengths (764, 784, 804, and 824 nm) were applied sequentially to obtain spectrally different PA images. Figure 21A and B shows the SO_2 (estimated with eq 6) and HbT (estimated with eq 7) mapping on the tumor-bearing brain in a nude mouse, respectively. The SO_2 in the tumor region is lower than the surrounding blood vessels indicating tumor hypoxia; the HbT is higher, indicating probable angiogenesis. SO_2 and HbT values from normal and tumor vasculatures were quantified in Figure 21C. The averaged SO_2 within the tumor is at least 13% lower than in the normal vasculature. In addition, the SO_2 level in the tumor is more heterogeneous as indicated by the error bars. Hence, a relative HbT (rHbT), defined as the ratio of the local HbT to the mean HbT in the medium fissure of each mouse, is adopted. Figure 21D show the plot




Figure 21. In vivo functional PA images of U87 glioblastoma xenograft in a nude mouse brain. (A) SO_2 image of the nude mouse brain (red arrow: hypoxic region). (B) HbT image of the nude mouse brain. (C) Comparison of normal and tumor vasculatures in SO_2 ($N = 3$). (D) The SO_2 versus the relative HbT in normal and tumor vasculatures. Reprinted with permission from ref 42. Copyright 2008 The Institute of Electrical and Electronics Engineers, Inc.

of SO_2 versus relative HbT (rHbT) between the normal and tumor vasculatures. The normal vessels tend to have a lower rHbT, a higher SO_2 , and a lower SO_2 fluctuation as compared to the tumor vessels.

3.5. Label-Free Functional Chronic Imaging of Wound Healing Process Using OR-PAM⁵⁰

Chronic imaging of long-term microhemodynamics plays an important role in the studies of disease progression,^{51,52} neural dynamics,⁵³ and functional recovery from pathological states.⁵⁴ Because the current microvascular imaging techniques require invasive procedures and/or fluorescence labeling, these effects disturb the normal physiology of the microcirculation, discourage chronic studies, and impede clinical translations. By avoiding invasiveness and phototoxicity, label-free OR-PAM (see section 2.2.3) is an ideal imaging technique for chronic microcirculation studies. The




Figure 22. In vivo noninvasive OR-PAM monitoring of the healing process of a laser-induced microvascular lesion. (A) Before laser treatment. (B) Immediately after laser treatment. (C) On each of the subsequent 12 days. Left column, photographs taken by a commercial transmission-mode optical microscope; middle column, the front view of the 3D microvascular morphology acquired by OR-PAM at 570 nm; right column, SO_2 mapping of the laser-damaged region overlaid with the morphological PA image. Reprinted with permission from ref 50. Copyright 2009 The American Association of Physicists in Medicine.

healing process of a laser-induced microvascular lesion in a nude mouse ear was monitored over a period of 12 days. At each stage of the chronic imaging, the corresponding transmission-mode optical microscopic photograph (left column in Figure 22) was acquired along with an OR-PAM image (middle column in Figure 22) at 570 nm. The SO_2 values (right column in Figure 22) within the microvascular lesion were estimated using two optical wavelengths, 570 nm (isosbestic point) and 578 nm (oxyhemoglobin dominant). Directly after acquiring the images in Figure 22A, a microvascular lesion in the nude mouse ear was created by 150 mW continuous wave (cw) laser illumination in the middle of the region of interest (ROI). This ROI was imaged immediately after the cw laser treatment (Figure 22B), and in the subsequent 12 days (Figure 22C, (C-1)–(C-12)) using both the OR-PAM and a conventional optical microscope. Four steps of the wound healing process were observed:⁵⁵ (1) vessel regression and hemostasis immediately following the laser treatment (Figure 22B); (2) inflammation (in the forms of vasodilation), which began 24 h after the injury and lasted for about 5 days (Figure 22C, (C-1)–(C-5)); hypoxia facilitated the synthesis of the VEGF to trigger angiogenesis (Figure 22B and C, (C-1)–(C-5)); (3) the growth of new capillaries, previously supplied by the damaged arteriole and observable after 3 days of the treatment (Figure 22C, (C-3)); and (4) after 12 days, the damaged arteriole–

venule pair was almost completely recovered (Figure 22C, (C-12)). The wound healing process could only partially be monitored by the commercial transmission-mode optical microscope.

4. Functional and Molecular PAT Using Exogenous Contrasts

Intrinsic optical contrasts, such as hemoglobin and melanin, are highly promising for PAT, but these contrasts significantly absorb light and generate strong PA signals in the visible spectral region. Because of both strong optical absorption and scattering, the light penetration depth is quite limited in this spectral region, and the maximum imaging depth is shallow. To image in deep tissues, the use of near-infrared (NIR) light is highly desirable for optical imaging. Exogenous contrast agents greatly enhance the sensitivity of PAT in the NIR spectral region. Additionally, some biological targets do not have intrinsic optical absorption contrast for PAT in the visible and near-infrared (NIR) regimes. One such important example is the lymphatic system. Through the use of contrast agents in PAT, imaging sensitivity and specificity are significantly improved just as in CT, PET, and MRI. Optically absorptive organic dyes, nanoparticles, reporter genes, and fluorescence proteins have been successfully applied as PA contrast agents.

4.1. Organic Dyes

4.1.1. Indocyanine Green (ICG)^{56,57}

Indocyanine green (ICG) is a nontoxic, water-soluble tricarbocyanine dye with a peak optical absorption at 790 nm.⁵⁸ The penetration depth of PA imaging using ICG can be increased using NIR light. The chemical structure of ICG is shown in Figure 23. ICG is FDA-approved for determining human cardiac output, hepatic function and blood flow, and ophthalmic angiography. Because ICG has a moderate fluorescence quantum yield ($\sim 10\%$ in DMSO and less than 1% in water),⁵⁹ it emits a sufficient amount of fluorescence and generates strong PA signals and as such can be used as a contrast agent for both fluorescence and PA imaging. Typical optical, chemical, and physical parameters of ICG are summarized in Table 1. The product of the molar extinction coefficient profile and the nonradiative quantum yield ($1 - \text{fluorescence quantum yield}$) of ICG is shown in Figure 31, where the effect of the internal conversion on the energy conversion is neglected. Because of aggregation of ICG molecules, the spectrum of ICG in water is highly unstable as the concentration is increased.⁶⁰ The spectrum changes are also caused by protein binding.

A single-element unfocused ultrasound transducer (10 MHz)-based PACT (see section 2.1.1) was employed to assess the contrast enhancement in brain cortical structures in Sprague–Dawley rats (~ 150 g) in vivo. Polyethylene glycol (PEG) was used to stabilize the ICG, thereby prolonging its circulation time in blood.⁶¹ This PEG conjugated ICG (ICG-PEG) highly absorbs light around an optical wavelength of 805 nm. The ICG-PEG (the concentration = $320 \mu\text{M}$) in phosphate buffered saline was injected intravenously through the tail vein. A dosage of $0.25 \text{ mL}/100 \text{ g}$ body weight was administered, and the estimated ICG concentration in blood was $\sim 10 \mu\text{M}$. Based on the injection dose, the number of detectable molecules in one resolved volume ($60 \mu\text{m} \times 60 \mu\text{m} \times 60 \mu\text{m}$ with the z dimension limited by the blood vessel diameter) is ~ 2 fmol per imaging voxel. Figure 24A and B shows the noninvasive PA images of a rat brain cortex before and after the ICG-PEG injection, respectively. The brain cortical structures of two PA images match well with the photograph (Figure 24D) taken after the PA imaging. The subtracted image (Figure 24C) between pre- and postinjection images shows increased PA signals due to the ICG injection. Moreover, this subtracted image




Figure 23. Chemical structure of ICG.⁵⁸




Figure 24. In vivo noninvasive PA images of a rat brain using ICG-PEG as contrast agents. (A and B) PA images obtained before and after the injection, respectively. MF, median fissure. (C) Subtracted PA image between before and after the injection ($C = B - A$). (D) Open-skull photograph after the PA imaging. Reprinted with permission from ref 57. Copyright 2004 The Optical Society of America.

shows a more detailed vasculature (indicated as black arrows in Figure 24C) than the images in Figure 24A and B.

To demonstrate the benefits of using contrast agents for deep PA imaging, the maximum imaging depth was increased to 5.2 cm in biological tissues, imaged ex vivo. Three pairs of transparent plastic tubes containing anticoagulation-treated whole rat blood, ICG solution in water (323 mM), and ICG solution in blood (129 mM) were embedded in chicken breast tissues of $\sim 7.5 \text{ cm}$ diameter in their transverse cross sections (Figure 25A). The optical coefficients of blood, ICG in water, and ICG in blood are ~ 4 – 5 , 16, and 43 cm^{-1} at 800 nm, respectively.⁶⁰ A single-element unfocused ultrasound transducer (2.25, 3.5, or 5 MHz)-based PACT (see section 2.1.1) was employed. Cross-sectional PA images were acquired by layering chicken breast tissues, so the imaging depth from the laser illumination surface was increased from 1.3 to 5.2 cm (Figure 25B). Figure 25C–F shows the PA images obtained at various depths (1.3, 2.6, 4.2, and 5.2 cm) using a 2.25 MHz transducer. The PA signals from the whole blood are not visible below 2.6 cm deep. However, the PA signals from ICG in water and ICG in blood are clearly discernible at depths reaching 5.2 cm. The PA signals from ICG in blood are stronger than those from ICG in water, which agree well with the above-

Table 1. Summaries of Chemical and Optical Properties of Organic Dyes

	dyes		
	Indocyanine green ⁵⁸	IRDye800-NHS ⁶²	Methylene blue ^{67,74}
molecular weight [g/mol]	775	1165	320
molecular formula	$\text{C}_{43}\text{H}_{47}\text{N}_2\text{NaO}_6\text{S}_2$	$\text{C}_{50}\text{H}_{54}\text{N}_3\text{Na}_3\text{O}_{17}\text{S}_4$	$\text{C}_{16}\text{H}_{18}\text{N}_3\text{ClS}$
peak molar extinction coefficient [$\text{cm}^{-1} \text{ M}^{-1}$]	1.5×10^5	2.7×10^5	7.4×10^4
peak absorption wavelength [nm]	790	795	667
fluorescence quantum yield [%]	10% in DMSO < 1% in water	15%	0.04%
peak molar extinction coefficient \times nonradiative quantum yield [$\text{cm}^{-1} \text{ M}^{-1}$]	1.4×10^5	2.3×10^5	7.1×10^4




Figure 25. Two-dimensional cross-sectional PA images at various depths in biological tissues. Photographs of (A) the cross section of chicken breast tissues where three pairs of tubes (whole blood, ICG in water, and ICG in blood) are embedded and (B) the entire sample. Two-dimensional PA images using a 2.25 MHz ultrasonic transducer at depths of (C) 1.3, (D) 2.6, (E) 4.2, and (F) 5.2 cm. Two-dimensional PA images at a 5.2 cm depth acquired by (G) 3.5 and (H) 5.0 MHz ultrasonic transducers. Reprinted with permission from ref 56. Copyright 2005 The Optical Society of America.

mentioned optical absorption coefficients of two solutions. Figure 25G and F shows the cross-sectional PA images at the depth of 5.2 cm using 3.5 and 5 MHz ultrasonic transducers, respectively. Despite lower SNRs at higher ultrasonic frequencies, the spatial resolution is better with the 5 MHz transducer.

4.1.2. IRDye-800 as Molecular Imaging Contrast Agents⁴²

ICG is hard to be functionalized for labeling. IRDye 800⁶² is a water-soluble ICG derivative dye for protein/antibody labeling applications. By adding the NHS ester reactive group, IRDye800-NHS is able to be functionalized for labeling primary and secondary amino groups. The chemical structure of IRDye800-NHS is shown in Figure 26. Typical optical, chemical, and physical parameters of IRDye800-NHS are summarized in Table 1. The product of the molar extinction coefficient profile and the nonradiative quantum yield of IRDye-NHS is shown in Figure 31.

A single-element unfocused ultrasound transducer (10 MHz)-based PACT (see section 2.1.1) was employed to image molecular contrast enhancement in nude mice (~20 g) in vivo using IRDye800-NHS as a contrast agent. Human U87 glioblastoma tumor cells (1×10^6 cells of $7 \mu\text{L}$) were inoculated intracranially into the caudate nucleus at a 3 mm depth from the skin surface.⁴⁹ IRDye800-NHS was conjugated with cyclic peptide cyclo(Lys-Arg-Gly-Asp-Phe),




Figure 26. Chemical structure of IRDye800-NHS.⁶²




Figure 27. In vitro binding test of IRDye800-c(KRGDf) to integrin $\alpha_v\beta_3$ overexpressed U87 tumor cells. (A) Fluorescence image of IRDye800-c(KRGDf) ($6 \mu\text{M}$) incubated U87 cells for 3 min. (B) Fluorescence image of U87 cells incubated with an excess 1.2 mM of c(KRGDf) for 5 min and then followed by a $6 \mu\text{M}$ of IRDye800-c(KRGDf) for 3 min. Red, IRDye800-c(KRGDf); green, Sytox green-labeled cell nuclei. Reprinted with permission from ref 42. Copyright 2008 The Institute of Electrical and Electronics Engineers, Inc.

which targets integrin $\alpha_v\beta_3$ (referred to as IRDye800-c(KRGDf)⁶³). Because IRDye800 absorbs light and emits fluorescence in the NIR regime, the former affects contrast in PACT and the latter affects contrast in fluorescence imaging, it can be used as a contrast agent for both imaging techniques. PAT was used to image the level of IRDye800-c(KRGDf) uptake in the brain tumor xenograft. First, before the injection, simultaneous SO_2 and HbT maps in the brain were acquired (see section 3.4). Conventional planar fluorescence imaging was employed to confirm the contrast-agent uptake. Integrin $\alpha_v\beta_3$ is overexpressed in newly formed tumor microvasculatures and important in progression, angiogenesis, and metastasis of melanoma and glioblastoma.⁶⁴ Figure 27 shows the binding capability of IRDye800-c(KRGDf) to U87 tumor cells in vitro.

Twenty hours after the intravenous injection of IRDye800-c(KRGDf), the mice were imaged first by a planar fluorescence imaging system and then by the PACT system. After the dual-modality imaging experiments, the brain was histologically analyzed to validate the results. The uptake of the IRDye800-c(KRGDf) in the tumor was first proven by the fluorescence imaging in Figure 28A. Figure 28B and C shows the in vivo molecular PACT image of IRDye800-c(KRGDf) and a composite image superimposing the segmented molecular PACT image (ultrasonic frequency, 2.25 MHz; in-plane spatial resolution, $312 \mu\text{m}$) onto the structural image (ultrasonic frequency, 10 MHz; in-plane spatial resolution, $60 \mu\text{m}$; optical wavelength, 804 nm) of the brain cortex, respectively. The uptake of the contrast agents is clearly seen in these PA images.




Figure 28. In vivo molecular PA images of a nude mouse brain with a U87 glioblastoma xenograft using IRDye800-c(KRGDf) as a molecular contrast agent. (A) A composite image of a fluorescence image and white-light photograph of the mouse head acquired 20 h after the injection. (B) In vivo molecular PA image of IRDye800-c(KRGDf). (C) A composite image of a segmented molecular and structural image, both acquired with the PACT system. Reprinted with permission from ref 42. Copyright 2008 The Institute of Electrical and Electronics Engineers, Inc.




Figure 29. Chemical structure of methylene blue (MB).⁶⁷

4.1.3. Methylene Blue (MB) as a Contrast Agent on Sentinel Lymph Node (SLN) Mapping⁶⁵

Methylene blue (MB)^{66,67} is a heterocyclic aromatic chemical compound widely used in biology and chemistry. In biology, MB is used as a stain in bacteriology and as an oxidation–reduction indicator.^{68,69} The chemical structure of MB is shown in Figure 29. Typical optical, chemical, and physical parameters of MB are summarized in Table 1. The product of the molar extinction coefficient profile and the nonradiative quantum yield of MB is shown in Figure 31.

Sentinel lymph node biopsy (SLNB), that is, the biopsy of the first lymph node receiving drainage from a cancer-containing area, has become the standard procedure for staging breast cancer patients to reduce the postoperative complications of axillary lymph node dissection (ALND).⁷⁰ Although SLNB with blue dye (lymphazurin blue or methylene blue) and radioactive tracers has an identification rate of 90–95% and a sensitivity of 88–95%, these methods still require ionizing imaging tools and intraoperative procedures. Moreover, they can fail to identify axillary disease due to a false negative rate of 5–10%.⁷¹ SLNB, despite being less invasive than ALND, still poses significant side effects.⁷² Recently, ultrasound-guided fine needle aspiration biopsy (FNAB) has been tested and clinically evaluated as a minimally invasive biopsy procedure.⁷³ However, this technique requires accurate identification of the SLN. Therefore, there is still a pressing need to develop an accurate, nonionizing, and noninvasive detection method for lymph node assessment. Thanks to the strong optical absorption coefficient of MB, PA imaging can be an alternative method to localize a SLN both accurately without the use of any ionizing radiation. In the United States, lymphazurin blue (LB)⁷⁴ is the only dye approved by FDA for SLN identification in breast surgery. However, LB causes following the complications: allergic and anaphylactic type I hypersensitivity reactions, skin pigmentation, and discoloration of body fluids. In addition, the supply of LB is insufficient, and LB is more expensive than MB. MB is more widely used in SLNB because it is more readily available and cost-effective. Furthermore, although MB causes more skin staining, no anaphylaxis has been reported for SLNB.




Figure 30. In vivo noninvasive PA images of the SLN in a rat. (A) Photograph with hair removed before PA imaging. (B) Photograph with skin removed after PA imaging. (C) Control image before the MB injection. (D) PA image immediately after the injection. (E) 52 min postinjection PA image with the scan head repositioned. (F) PA image of the deeply positioned (~18 mm) SLN. Reprinted with permission from ref 65. Copyright 2008 Society of Photo-Optical Instrumentation Engineers.

A deep reflection-mode PAM system (see section 2.2.2) was employed to map SLNs noninvasively in Sprague–Dawley rats (~250–350 g) using MB. An ultrasonic frequency of 5 MHz and an optical wavelength of 635 nm were used in this system. The incident laser fluence on the rat skin was 3.4 mJ/cm², within the ANSI safety limit (20 mJ/cm²).²² After hairs were removed from the rat skin, a control image was acquired. 0.07 mL of 1% MB (10 mg/mL) was intradermally injected on a left forepaw pad. After the administration of MB, a series of PA images were acquired. Figure 30A and B shows photographs of a rat taken before and after image acquisition, respectively. After PA imaging, the rat's skin was removed to identify the MB dyed lymph node (Figure 30B). The surrounding vasculature in the axillary area is clearly seen with an image contrast of 7 ± 0.5 with respect to the background in the control image (Figure 30C). The SLN appeared immediately after the injection of MB with an image contrast of 12 ± 2 , seen in Figure 30D. The SLN was still observed 50 min postinjection after the scan head was repositioned (Figure 30E). The mean depth of the top surface of human SLNs is 12 ± 5 mm in ultrasonography. To determine the feasibility of PAT in human SLN imaging, the imaging depth was increased to 18 mm by adding chicken tissues on top of the rat. Although the blood vessels faded away, the SLN is still clearly seen in Figure 30F.

Table 1 summarizes the optical, chemical, and physical parameters of organic dyes. The products of the molar extinction coefficients and nonradiative quantum yield values, defined as the difference of the fluorescence quantum yield from 1, of various organic dyes are shown in Figure 31. The molar extinction coefficients of oxy (HbO₂) and deoxyhemoglobin (Hb) are also plotted in the same figure.




Figure 31. Molar extinction profiles of oxyhemoglobin (HbO₂), deoxyhemoglobin (Hb), and organic dyes multiplied by nonradiative quantum yield.

4.2. Nanoparticles

Nanoparticles have received considerable attention in biomedical applications, because it has been proven that they can be utilized effectively for delivering^{75,76} and targeting^{77,78} therapeutic agents to a specific site and as contrast agents for diagnostic purposes.^{78–80} Because of their tunable properties such as sizes, shapes, and compositions, nanoparticles are widely used to enhance image contrasts throughout biomedical imaging fields, including optical imaging,^{78,81,82} magnetic resonance imaging (MRI),^{83–85} and X-ray computed tomography (CT).^{86,87} The basic theories and fabrication processes of gold nanostructures (gold nanospheres, gold nanoshells, gold nanorods, and gold nanocages) used in PAT were reviewed in ref 88. In this Review, the biomedical applications of gold nanostructures in conjunction with PAT will be reviewed.

4.2.1. Gold Nanocages^{89,90}

Gold nanocages are hollow porous nanoparticles with a broad range of sizes (35–100 nm). Gold nanocages have been used as a contrast agent in PA imaging because of biocompatibility, easily modifiable surfaces for targeting,^{91,92} a lack of heavy metal toxicity,⁹³ a tunable localized surface plasmon resonance (LSPR) peak, encapsulated site-specific drug delivery,⁹⁴ and strong optical absorption in the NIR spectral region.^{38,91}

Gold nanocages were prepared by titrating silver nanocubes with aqueous HAuCl₄.^{94–100} Silver nanocubes with an edge length of ~30 nm (Figure 32A) were synthesized using a polymer-mediated polyol method, which involves reduction of silver nitrate by ethylene glycol at 160 °C.^{101–103} In general, the size of the nanocubes could be varied by controlling the reaction time. The silver nanocubes can then serve as sacrificial templates and be converted into gold nanocages via a galvanic replacement reaction, $3\text{Ag(s)} + \text{AuCl}_4^-(\text{aq}) \rightarrow \text{Au(s)} + 3\text{AgCl(s)} + \text{Cl}^-(\text{aq})$, with HAuCl₄ in water under refluxing (Figure 32B). The edge length of the subsequent gold nanostructure increased from 30 to 36 nm because the gold was deposited around the outer surface of the silver template. The ensuing gold nanocages have a strong absorption peak, determined by the ratio of the wall thickness to edge length. Experimentally, we can easily and precisely control this parameter by varying the molar ratio of Ag to HAuCl₄ in a fashion similar to acid–base titration. Figure 32C shows the extinction spectra measured for the 30 nm silver nanocubes and the resultant gold nanocages with the extinction peak specifically tuned to 800 nm. Using




Figure 32. (A) Transmission electron microscopy (TEM) image of 30 nm silver nanocubes synthesized with a growth time of 20 h. (B) Scanning electron microscopy (SEM) image of gold nanocages prepared with the silver cubes as a sacrificial template. The inset shows a TEM image of the same sample. (C) Extinction spectra taken from aqueous suspensions of the silver nanocubes and the corresponding gold nanocages. (D) Spectra calculated using the DDA method for a gold nanocage with the dimensions shown in the inset (extinction = absorption + scattering). Reprinted with permission from refs 91 and 104. Copyright 2005 American Chemical Society.




Figure 33. (A) Schematic illustration of the formation of immuno gold nanocages. (B) SEM image of a typical SK-BR-3 cancer cell whose surface was directly targeted by immuno gold nanocages of 36 nm in size. The inset shows an image at higher magnification (scale bar = 100 nm). (C) SEM image of a typical SK-BR-3 cell that had been incubated with PEGylated nanocages (i.e., no specific targeting), showing essentially no nanocages on the surface. Reprinted with permission from refs 91 and 104. Copyright 2005 American Chemical Society.

the discrete dipole approximation (DDA) method, we can predict the exact scattering and absorption components (Figure 32D) for the gold nanocages. The calculated and measured spectra agree well in terms of peak position. Calculations indicate that absorption is predominant for gold nanocages <50 nm in size.¹⁰⁴

Using established alkanethiol chemistry, the surface of the gold nanocages can be easily functionalized with tumor-targeting moieties such as antibodies or polypeptides.¹⁰⁴ For instance, the surface of gold nanocages can be derivative with thiol molecules containing the poly ethylene-glycol (PEG) segment and *N*-hydroxysulfosuccinimide (NHS) terminus, a facile leaving group for amide bond formation. The NHS terminus can be coupled into the antibodies (Figure 33A). To confirm the effectiveness of bioconjugation, the targeting specificity was studied *in vitro*. As an example, a

breast cancer cell line (SK-BR-3), which is known to overexpress HER2 receptor on its surface, was chosen for demonstration. Monoclonal antibodies (i.e., anti-HER2) were employed to target the receptors. Figure 33B shows a scanning electron microscopy (SEM) image taken from a representative cancer cell whose surface was directly targeted with anti-HER2-derivatized gold nanocages, revealing the molecular specificity and high coverage of nanocages on the cell surface. As a control, the SK-BR-3 cells were incubated with PEGylated gold nanocages that did not have anti-HER2 attached. The SEM image (Figure 33C) shows essentially no gold nanocages on the cell surface. These results demonstrate that the immuno gold nanocages can selectively target the specific receptor expressed on the cancer cell surface.

A single-element unfocused ultrasound transducer (10 MHz)-based PACT (see section 2.1.1) was employed to confirm the contrast enhancement of gold nanocages for PA imaging in vivo. PEGylated gold nanocages were injected into Sprague–Dawley rats (100–150 g) via the tail vein, and PA images of its cerebral cortex were recorded. Three successive injections of nanocages were administered, each with a dose of $\sim 0.8 \times 10^9$ nanocages/g body weight. The data were acquired before and after injection up to 3 h with a optical wavelength of 804 nm. Figure 34A shows the structural image of the rat brain, measured from intrinsic optical contrast, before injection of PEGylated nanocages, while Figure 34B shows a better resolved image of the same rat brain after the injection. Figure 34C shows the subtracted image between the preinjection image (Figure 34A) and the postinjection image (Figure 34B). All PA images of the rat brain cortex (Figure 34A–C) match well with the open-skull anatomical photograph in Figure 34D. The enhancement of PA signals due to gold nanocages was quantified as a function of time in Figure 34E. At 2 h after the final injection, the PA signal was increased maximally by $\sim 81\%$, after which the signal gradually cleared out from the blood.

In another demonstration, 100 μL of 2 nM gold nanocage solutions was intradermally injected on the left forepaw pads of Sprague–Dawley rats (250–350 g) to image sentinel lymph nodes noninvasively. The deep reflection-mode PAM system (see section 2.2.2) was employed. Figure 35A shows a photograph of the axilla of a rat after the hair was removed and before PA imaging. The photograph in Figure 35B shows the same region of the rat with the skin removed and the SLN containing gold nanocages after the PA imaging. The control image is shown in Figure 35C, captured prior to the injection of gold nanocages. The surrounding vasculature is clearly seen in the control image. The SLN was detectable about 5 min after the injection (Figure 35D). Figure 35E–G shows the time-course of PA images acquired at 54, 135, and 189 min after the injection, respectively. Figure 35H shows the increase of PA signal in the SLN as a function of time. As time elapsed, the gold nanocages gradually accumulated, and the PA signal within the SLN was increased. The peak accumulation time of gold nanocages in the SLN was observed to be about 140 min after the injection.

4.2.2. Gold Nanoshells^{82,105,106}

Gold nanoshells comprised of silica cores (~ 116 nm in diameter) coated with a thin gold shell (~ 14 nm) have widely tunable plasmon resonances in the visible and near-infrared regimes.¹⁰⁷ The absorption peaks of these nanoshells are tunable by varying the core size relative to the thickness




Figure 34. In vivo noninvasive PA images of the cerebral cortex acquired (A) before the injection and (B) ~ 2 h after the final of PEGylated nanocages via tail vein. (C) A differential image ((B) subtracted from (A)). (D) An open-skull photograph of the rat's cerebral cortex. (E) Quantification of the PA signal enhancement after the injection as a function of time. Reprinted with permission from ref 90. Copyright 2007 American Chemical Society.

of the gold shell. In addition, the optical absorption properties of the gold nanoshells also depend on rigid metallic structures, unlike organic dyes, whose optical properties are determined by molecular orbital electronic transitions. Because of their excellent biocompatibility, gold nanoshells have been widely used in photothermal therapy¹⁰⁸ and as diagnostic contrast agents.^{82,105} PEGylated gold nanoshells have longer blood circulation time.

A single-element unfocused ultrasound transducer (10 MHz)-based PACT (see section 2.1.1) was employed to confirm the contrast enhancement of gold nanoshells for PA imaging in vivo. The extinction spectrum of the gold nanoshells peaked at 800 nm, and accordingly an 800 nm laser excitation wavelength was used in PA imaging. PEGylated gold nanoshells were injected into Sprague–Dawley rats (130–160 g) via the tail vein, and then a series of PA images were acquired of their cerebral cortices. The nanoshells were administered three times successively. Each dose was $\sim 0.8 \times 10^9$ nanoshells/g body weight. After the final injection, the rat brain was imaged sequentially 10 times for more than




Figure 35. Noninvasive in vivo PA images of the SLN in a rat. (A) Photograph of axillary region with hair removed before the injection of gold nanocages. (B) Photograph with skin removed after the PA images had been captured. PA images acquired before (C) and after (D–H) the nanocage injection: (D) 5 min (SLN started to appear), (E) 59 min, (F) 140 min, (G) 194 min. (H) The increase of PA signal in the SLN as a function of time. BV, blood vessels; SLN, sentinel lymph node. Reprinted with permission from ref 89. Copyright 2009 American Chemical Society.

~6 h. Figure 36A shows the brain cortical vasculature before injection of PEGylated nanoshells. A PA image of rat brain cortex is shown in Figure 36B, acquired ~20 min after the injection. Figure 36C shows the subtracted image between the preinjection image (Figure 36A) and the postinjection image (Figure 36B). This image shows the distribution of increased optical absorption in the rat brain due to the injection of PEGylated nanoshells. All PA images of the rat cerebral cortex (Figure 36A–C) match well with the open-skull anatomical photograph in Figure 36D taken after the PA imaging. The time-course PA signal increase was quantified in Figure 36E. The maximal increase of PA signal was ~63%. After that, the signal gradually decreased.

In another demonstration, the 50 MHz dark-field confocal PAM (see section 2.2.1) was employed to image progressive extravasation and accumulation of nanoshells within a solid tumor in vivo. Immunocompetent BALB/c mice (~20 g) were inoculated with CT26.wt murine colon carcinoma cells subcutaneously on the mice brains. Imaging was performed 7 days after tumor inoculation. PEGylated nanoshells with a peak optical absorption at 800 nm were then intravenously




Figure 36. In vivo noninvasive PA images of the cerebral cortex acquired (A) before the injection and (B) ~20 min after the final injection of PEGylated nanoshells via tail vein. (C) A differential image ((B) subtracted from (A)). (D) An open-skull photograph of the rat's cerebral cortex showing that the PA images agreed with the location of blood vessels. (E) Quantification of the PA signal enhancement after the injection as a function of time. Reprinted with permission from ref 82. Copyright 2004 American Chemical Society.

administered at a dose of 1.4×10^9 nanoshells/g body weight. A control imaged prior to the nanoshell injection was acquired in Figure 37A. After the injection, the extravasation and accumulation of nanoshells in and around the tumor was monitored for 6 h. Figure 37B–D shows the passive uptake of nanoshells at the tumor at 1.4, 3, and 5.8 h postinjection, respectively. As time elapsed, the accumulation of nanoshells at the tumor is clearly evident. Time-dependent PA signal enhancement in the vessels and tumor from the injection is shown in Figure 37E. At 5.8 h postinjection, the image contrast ratio of tumor to vessels is ~6.5.

A single-element focused ultrasound-based PAM (48 MHz)¹⁰⁶ was employed to investigate the feasibility of molecular PA imaging of specific tumors using targeted spherical gold nanoparticles. Three tissue phantoms with human epithelial carcinoma cells (A431 keratinocyte) were tested: (1) a control phantom without any nanoparticles; (2) a targeted phantom with EGFR (epithelial growth factor receptor) targeted gold nanoparticles; and (3) a nontargeted phantom with PEGylated gold nanoparticles. The absorbance spectra of the control, targeted, and nontargeted phantoms are shown in Figure 38A. The nontargeted phantom has a peak absorbance at 520 nm, while the peak absorbance of




Figure 37. In vivo noninvasive PA images of nanoshell extravasation in a tumor. (A) Control image before the injection of nanoshells. (B, C, and D) PA images of accumulation of nanoshells at 1.4, 3, and 5.8 h postinjection, respectively. (E) Averaged PA signals as a function of time. Line with circles, vessel regions; dashed line with squares, tumor foci. Reprinted with permission from ref 105. Copyright 2009 Society of Photo-Optical Instrumentation Engineers.

the targeted phantom has red-shifted and broadened due to EGFR-mediated aggregation of gold nanoparticles.¹⁰⁹ In the control phantom, the absorbance is significantly lower throughout a wide spectral range. On the basis of Figure 38A, the PA images of three phantoms shown in Figure 38B, D, and F were acquired at the optical wavelength of 532 nm. The other three images, Figure 38C, E, and G, were obtained at 680 nm. Nothing is shown in the control images at both wavelengths in Figure 38B and C, which agrees well with the absorbance spectrum of the control phantom. Stronger PA signals are observable from the nontargeted phantom (Figure 38F) than from the targeted phantom (Figure 38D) at 532 nm, whereas opposite results were recovered when illuminated with an optical wavelength of 680 nm. These results agree well with the absorbance spectra in Figure 38A.

4.2.3. Gold Nanorods^{18,110–113}

Because gold nanorods possess dielectric properties in the NIR, they have a strong and narrow optical absorption band in this regime. Additionally, gold nanorods have biocompatibility, surfaces that are easily modifiable and amenable for targeting, a lack of heavy metal toxicity, and a readily tunable LSPR peak. Because of these properties, gold nanorods have been widely used in photothermal therapy¹¹⁴ and as diagnostic contrast agents in PAT.^{18,110–113}




Figure 38. (A) Absorbance spectra of control, targeted, and nontargeted phantoms. (B, D, and F) PA images ($\lambda = 532$ nm) of control, targeted, and nontargeted phantoms, respectively. (C, E, and G) PA images ($\lambda = 680$ nm) of control, targeted, and nontargeted phantoms, respectively. Reprinted with permission from ref 106. Copyright 2007 The Optical Society of America.

A single-element focused ultrasound-based PAM (50 MHz)¹¹² was employed to demonstrate the feasibility of PA imaging of inflammatory responses using bioconjugated gold nanorods in vitro. The aspect ratio of gold nanorods was 3:1, and the peak wavelength of optical absorption was at 700 nm. Intercellular adhesion molecule-1 (ICAM-1) is a strong inflammatory mediator in systemic inflammatory diseases such as rheumatoid arthritis (RA).¹¹⁵ Anti-intercellular adhesion molecule-1 (anti-ICAM-1) was selected as a biomarker of inflammation and activation of endothelial cells (ECs), which binds to cell surfaces overexpressing ICAM-1. Unstimulated and stimulated ECs were incubated with gold nanorods bioconjugated with anti-ICAM-1, photoacoustically imaged, and compared. Nonbioconjugated gold nanorods (blank gold nanorods) incubated in both unstimulated and stimulated ECs. 10^{12} nanorods/mL were used as a control for all experiments. PA signals from stimulated ECs with bioconjugated nanorods (Figure 39A) were 10 dB stronger than those from unstimulated ECs with bioconjugated nanorods (Figure 39B). This result implies ICAM-1 was overexpressed on the surfaces of stimulated ECs, and anti-ICAM-1 conjugated nanorods were highly bound to the cell surfaces. When blank nanorods were used, PA signals from both stimulated and unstimulated ECs did not increase significantly, as seen in Figure 39C and D, respectively. This reflects very low nonspecific binding efficiency of blank nanorods to ECs.




Figure 39. PA images of (A) stimulated and (B) unstimulated ECs binding with anti-ICAM-1 conjugated gold nanorods, and (C) stimulated and (D) unstimulated ECs binding with blank gold nanorods. Reprinted with permission from ref 112. Copyright 2007 American Institute of Physics.

The custom-made ultrasonic array-based PA imaging system¹¹¹ was employed to investigate the feasibility of preclinical and clinical applications of PA imaging using gold nanorods as a contrast agent. A nude mouse was positioned on top of the ultrasonic array system, and light illuminated from the top. The PA imaging system consists of two parallel linear arrays of 32 acoustic transducers equipped with a real-time multichannel signal processor. The excitation laser wavelength of 757 nm was close to the peak wavelength of optical absorption of the nanorods. Using a single channel acoustic transducer, the minimum sensitivity was measured in nude mice. First, PBS (25 μ L) was injected subcutaneously into the abdominal area of the mouse, and a control PA image was acquired. Next, PEGylated gold nanorods with a concentration of 1.25 pM (7.5×10^8 nanorods/mL) were injected subcutaneously. The localized PEGylated gold nanorods are clearly seen in the subcutaneous (SQ) tissue, where the gold nanorods were injected, in Figure 40A. Using the ultrasonic array-based PA imaging system, the PA signal enhancement due to gold nanorods was observed in a nude mouse in vivo. PEGylated gold nanorods of 100 μ L with a concentration of 7.5×10^{10} nanorods/mL were subcutaneously injected into the abdominal area of the mouse. Figure 40B and C shows two PA images before and after the nanorod injection, respectively. The injected gold nanorods are clearly seen in Figure 40C with an image contrast of 9:1 as compared to the background.

A custom-made single-element ultrasound transducer-based PAM (20 MHz)¹⁸ was employed to demonstrate the feasibility of molecular PA imaging of malignant tumors using bioconjugated gold nanorods (AuNRs) in vivo. Two types of tumor cell lines were used in this study: (1) oral cancer OECM1 (oral squamous cell carcinoma) cells overexpressed with HER2 (human epidermal growth factor receptor 2) on the cell surface, and (2) Cal27 (squamous cell carcinoma) cells overexpressed with EGFR (epidermal growth factor receptor). For multiple targeting, two types of gold nanorods were tested: (1) AuNR₇₈₅-HER2 (peak absorption wavelength, 785 nm; aspect ratio, 3.7; conjugated antibody, anti-HER2 for targeting OECM1 cells) and (2) AuNR₁₀₀₀-EGFR (peak absorption wavelength, 1000 nm; aspect ratio, 5.9; conjugated antibody, anti-EGFR for targeting Cal27 cells). Two optical wavelengths, 800 and 1064 nm, were used for AuNR₇₈₅-HER2 and AuNR₁₀₀₀-EGFR, respectively. The AuNRs were coated with poly(ethylene




Figure 40. (A) In vivo noninvasive PA A-line image of a nude mouse after the injection of gold nanorods (25 μ L at a concentration of 1.25 pM) or PBS (25 μ L). The x -axis of (A) is the time of arrival of ultrasound. A typical PA image of a nude mouse before (B) and after (C) subcutaneous injection of 100 μ L of gold nanorods at a concentration of 7.5×10^{10} nanorods/mL in the abdominal area. Injected nanoparticles were brightly visible in the PA image (C). Reprinted with permission from ref 111. Copyright 2007 American Chemical Society.

glycol) (PEG). The PEG-AuNRs (referred to as AuNRs) without any bioconjugation were used for control experiments. 100 μ L of AuNRs was injected to mice in each study at concentration of 30 nM. Figure 41 shows time-course coregistered ultrasonic (US, gray color scale) and PA (red color scale) images of Cal27 tumors before and after the injections of AuNR₁₀₀₀ and AuNR₁₀₀₀-EGFR, independently. The contoured area indicates the tumor region in Figure 41A and B. Within the tumor region, the PA image contrast with the injection of AuNR₁₀₀₀-EGFR at 7 h postinjection is higher than the contrast prior to the injection (Figure 41A). When AuNR₁₀₀₀ was injected, there is no apparent difference in PA images between pre- and postinjections (Figure 41B). The time-dependent PA signal enhancement within the tumor region was quantified in Figure 41C. The contrast using AuNR₁₀₀₀-EGFR is maximally ~ 3.5 dB higher than that using AuNR₁₀₀₀ at 7 h postinjection.

Figure 42 shows time-course coregistered ultrasonic (US, gray scale) and PA (red scale) images of OECM1 tumors before and after the injections of AuNR₇₈₅ and AuNR₇₈₅-HER2, independently. Within the tumor region,




Figure 41. In vivo noninvasive molecular PA images of Cal27 tumors before and after the injection of AuNR₁₀₀₀-EGFR and AuNR₁₀₀₀. Ellipses indicate the tumor regions. (A) Time-course composite of US and PA images before and after AuNR₁₀₀₀-EGFR injection. US image in gray scale and PA image in red pseudocolor. (B) Time-course composite of US and PA images before and after AuNR₁₀₀₀ injection. (C) Averaged image intensities within the tumor region as a function of time after injections of AuNR₁₀₀₀-EGFR (blue solid line) and AuNR₁₀₀₀ (red dashed line). Reprinted with permission from ref 18. Copyright 2008 The Optical Society of America.

the PA image contrast obviously increased 1–17 h postinjection of AuNR₇₈₅-HER2 as compared to the contrast prior to the injection (Figure 42A). When AuNR₇₈₅ was injected, there is no apparent difference in PA images between pre- and postinjections (Figure 42B). The time-dependent PA signal enhancement within the tumor region was quantified in Figure 42C. The contrast using AuNR₇₈₅-HER2 is maximally ~2 dB higher than that using AuNR₇₈₅ at 14 h postinjection.

In another demonstration, 100 μ L of a 979-pM gold nanorods (AuNRs) solution was injected intradermally in the left forepaw pads of Sprague–Dawley rats (~400 g) to noninvasively image sentinel lymph nodes in vivo. The deep reflection-mode PA imaging system (see section 2.2.2) was employed. The AuNRs used for this study had a 10 nm




Figure 42. In vivo noninvasive molecular PA images of OECM1 tumors before and after the injection of AuNR₇₈₅-HER2 and AuNR₇₈₅. Ellipses indicate the tumor regions. (A) Time-course composite of US and PA images before and after AuNR₇₈₅-HER2 injection. US image in gray scale and PA image in red pseudocolor. (B) Time-course composite of US and PA images before and after AuNR₇₈₅ injection. (C) Averaged image intensities within the tumor region as a function of time after injections of AuNR₇₈₅-HER2 (blue solid line) and AuNR₇₈₅ (red dashed line). Reprinted with permission from ref 18. Copyright 2008 The Optical Society of America.

diameter, a 41 nm length, and a peak absorption wavelength of 807 nm. Figure 43A shows a photograph of the axilla of the rat after the hair was removed before PA imaging. Figure 43B shows a photograph of the same region with skin removed and shows the SLN containing AuNRs after the PA imaging, and the top and bottom views of the excised SLN. The control image is shown in Figure 43D captured prior to the injection of AuNRs. The surrounding vasculature is clearly seen in the control image. The SLN was not detectable until 19 h postinjection (Figure 43E and F). At 46 h postinjection, the SLN was clearly seen with good contrast, shown in Figure 43G. The depth-resolved B-scan along the dotted line in Figure 43G shows the depth information of SLN and blood vessels in Figure 43C.




Figure 43. In vivo noninvasive PA images of the axilla, showing the vasculature and sentinel lymph node. (A) Photograph of the axilla with hair removed before imaging. (B) Photograph of the axilla with tissue removed. SLN is indicated by an arrow. Insets are a top view and a bottom view of the excised SLN, respectively. (C) B-scan acquired along the dotted line in (G). (D) PA image taken before the injection as a control. (E) PA image at 1.5 h postinjection. (F) PA image at 19 h injection. (G) PA image acquired at 46 h postinjection. Reprinted with permission from ref 113. Copyright 2009 Elsevier.

4.2.4. Single-Walled Carbon Nanotubes^{19,116–118}

Although the toxicity of carbon nanotubes is still an ongoing debate in biomedical applications,^{119–121} single-walled carbon nanotubes (SWNTs) were tested in PA imaging as a contrast agent for molecular specific tumor targeting^{19,118} and sentinel lymph node mapping in vivo.¹¹⁶ SWNTs were also investigated as agents for photothermal therapy.¹²² The synthesis of SWNTs used in PA imaging is explained in detail in refs 116, 117.

A single-element focused ultrasound transducer-based PAM system (5 MHz)¹⁹ was employed to investigate the feasibility of molecular PA imaging of malignant tumors using bioconjugated SWNTs in vivo. SWNTs (1–2 nm in diameter and 50–300 nm in length) were coupled to the RGD peptides¹²³ (SWNT-RGD) through polyethylene glycol-5000 grafted phospholipid (PL-PEG5000) (Figure 44A). The SWNT-RGD targets to $\alpha_v\beta_3$ integrin overexpressed in tumor neovasculature. For control experiments, only PEGylated SWNTs (Plain SWNT) were synthesized without any biomarkers. Figure 44B shows the optical absorption profiles of oxy-hemoglobin (HbO₂), deoxy-hemoglobin (Hb), plain-SWNT, and SWNT-RGD. These profiles suggest that 690 nm is preferable for PA imaging. The PA sensitivity measured using SWNTs in biological tissues is 50 nM; that is, the PA signal from tissues (background) was equal to the PA signal from 50 nM of SWNTs (SNR = 1).

200 μ L of SWNT-RGD at a concentration of 1.2 mM was intravenously injected via the tail-vein to U87MG tumor xenografts bearing mice ($n = 4$). The same amount and concentration of plain-SWNT was intravenously injected as well ($n = 4$). A series of US and PA images around the tumor bearing region were acquired before and up to 4 h after injection. Figure 45A shows the time-dependent increase of PA signals in mice after the injection of SWNT-RGD and plain-SWNT. The PA images of mice injected with SWNT-RGD show significant signal enhancement within the tumor region. A differential image between the PA images taken at 4 h postinjection and the image taken before injection clearly proves the accumulation of SWNT-RGD within the tumor, as compared to the differential image using plain-SWNT. The PA signal increase was quantified as a function of time (Figure 45B). On average, the PA signals




Figure 44. (A) Illustration of plain-SWNT and SWNT-RGD. (B) Photoacoustic spectra of plain-SWNT and SWNT-RGD overlaid on optical spectra of oxyhemoglobin (HbO₂) and deoxyhemoglobin (Hb). Reprinted with permission from ref 19. Copyright 2008 Nature.

using SWNT-RGD are ~ 8 times greater than those using plain-SWNT at 4 h postinjection.

In another demonstration, 75 μ L of a 500 nM SWNTs solution was intradermally injected in the left forepaw pads of Sprague–Dawley rats (~ 250 – 350 g) to image sentinel lymph nodes in vivo noninvasively. The deep reflection-mode PA imaging system (see section 2.2.2) was employed. The optical wavelength of 807 nm was utilized. The control image




Figure 45. In vivo noninvasive molecular PA images of U87MG tumors before and after the injection of SWNTs-RGD and plain-SWNTs. The structural ultrasound image is shown in gray scale, and the PA image is shown in green scale. (A) Time-course composite of US and PA images before and after SWNT-RGD and plain-SWNT. Subtraction image: PA image at 4 h postinjection – control PA image. White arrows in the left column indicate the PA signals in the mouse injected with plain-SWNT. (B) Averaged image intensities within the tumor region as a function of time after injection of SWNTs-RGD ($n = 4$, red solid line) and plain-SWNTs ($n = 4$, blue solid line). Reprinted with permission from ref 19. Copyright 2008 Nature.

captured prior to the injection of SWNTs is shown in Figure 46A. The surrounding vasculature is clearly seen in the control image. The SLN appeared immediately after the injection (Figure 46B). Figure 46C–E shows the PA images of the same area 30, 55, and 85 min postinjection of SWNTs. The uptake kinetics of the SWNTs is shown in Figure 46F.

4.2.5. Colloidal Gold Nanobeacons¹²⁴

Colloidal gold nanobeacons (GNBs) are soft, hollow colloidal particles containing smaller gold nanoparticles (2–4 nm) inside. Based on the high optical absorption in the visible and NIR regimes, GNBs can be used as molecular contrast agents in PA imaging. The GNBs have a nominal hydrodynamic diameter of 154 ± 10 nm (Figure 47). The benefit of the bigger size of GNBs than other gold nanostructures (<50 nm) is that the GNBs have a greater chance to stay in the bloodstream and lead to specific targeting. Smaller particles can rapidly distribute beyond the vasculature and into tissues and bind to nonspecific targets, consequently causing an unwanted signal increase in background. The fabrication process of GNBs is described in detail in ref 124.




Figure 46. In vivo noninvasive PA images of the axilla, showing the vasculature and sentinel lymph node. (A) The control PA image taken before the injection. (B) PA image taken immediate after the injection. (C–E) PA images at 30, 55, and 85 min postinjection, respectively. (F) The uptake kinetics of the SWNTs. Reprinted with permission from ref 116. Copyright 2009 Institute of Physics.




Figure 47. Preparation of gold nanobeacons from octanethiol-functionalized gold nanoparticles (AuNPs). Reprinted with permission from ref 124. Copyright 2009 Wiley-VCH Verlag GmbH & Co.

A solution of GNBs (3 mL/kg) was intravenously injected to a Sprague–Dawley rat to preliminarily test noninvasive in vivo blood-vessel imaging of a femoral vein. The deep reflection-mode PAM system (see section 2.2.2) was employed, using an optical wavelength of 766 nm. The control image captured prior to the injection of GNBs is shown in Figure 48A. The vasculature is clearly seen in the control image with a contrast-noise ratio (CNR) of 50. Figure 48B shows the PA image (CNR = 68) of the same area 156 min after the injection. The PA signal in blood vessels increased up to 60% after injection as compared to that in control image. Figure 48C shows the time-course PA signal enhancement after the injection.

4.2.6. Nanowontons¹²⁵

Composite-material nanoparticles, “nanowontons”, were recently introduced as contrast agents for MRI and PAT. The nanowontons have a cobalt (Co) core for MRI and gold (Au) thin film coating. The nanowontons exhibit both ferromagnetic and optical responses, making them useful for dual-modality MRI and PAT studies. The shape and thickness of the Au capping layer are designed to match the peak optical absorption wavelength in the NIR (700 nm) regime. The fabrication process




Figure 48. In vivo noninvasive PA vascular structural images of a rat. (A) Control PA image acquired before GNB injection. (B) PA image acquired 156 min after injection of GNBs. (C) Pharmacokinetics of GNBs in a blood vessel after injection (blue arrow: point of GNBs injection). Reprinted with permission from ref 124. Copyright 2009 Wiley InterScience.




Figure 49. PA image of nanowonton phantom. (A) Optical absorbance of nanowontons. (B) PA image of four absorbing objects containing nanowonton concentrations of 100, 50, 25, and 13 pM, respectively, for objects A, B, C, and D. Reprinted with permission from ref 125. Copyright 2009 National Academy of Sciences.

of nanowontons is explained in detail in ref 125. The optical absorbance of nanowontons is shown in Figure 49A. Figure 49B shows the PA images of a tissue phantom containing four optically absorptive targets at various concentrations of nanowontons: 100, 50, 25, and 13 pM. The detection sensitivity of this system is on the order of 25 pM.

4.2.7. Quantification of Optical Absorption Cross Sections of Gold Nanostructures Using Photoacoustic Sensing¹²⁶

Gold nanostructures can have a strong extinction peak (localized surface plasmon resonance) in the visible and near-

infrared (NIR) regimes. This extinction peak is comprised of two components: scattering and absorption. By changing the size, shape, and internal structure of the nanostructures, the relative magnitude of light being absorbed and scattered can be tuned. In general, different applications may require nanostructures to possess very specific optical properties. For example, optical scattering dominant nanostructures can be used as a contrast agent in optical coherence tomography, and optical absorbing dominant nanostructures are more valuable to PA imaging. Therefore, it is essential to know both the absorption (σ_a) and the scattering (σ_s) cross sections of a nanostructure for a specific biomedical application.

A conventional UV–vis–NIR spectrometer measures the extinction spectra of nanostructures, which eventually provides the extinction cross-section (σ_e , with $\sigma_e = \sigma_a + \sigma_s$) of the nanostructures. Currently, theoretical simulation is the dominant method to separate σ_e into σ_a and σ_s . The Mie theory has been used for spherical particles, while the discrete-dipole approximation (DDA) is for other shapes of nanostructures.

The σ_a of gold nanostructures can be experimentally measured using PA sensing. The PA signal is primarily sensitive to optical absorption, and the signal is directly proportional to the absorption coefficient (μ_a) of the material. Therefore, PA signals from nanostructures at various concentrations can be converted into the corresponding μ_a by comparing it against a known linear calibration curve that (e.g., methylene blue or indocyanine green) describes the relationship between the PA signal and the μ_a of an organic dye. Next, σ_a can be obtained by dividing μ_a by each corresponding concentration. When combined with the conventional UV–vis–NIR spectroscopic measurement, the σ_e , σ_a , and σ_s of the nanostructure can be determined, independently.

Table 2 shows that the ratios of absorption to extinction cross sections (σ_a/σ_e) obtained from experimental results, which agree well with theoretical calculations. This agreement demonstrates that the PA sensing-based technique can be used to measure the absorption and scattering cross sections of gold nanostructures.

4.3. Reporter Genes^{127,128}

In gene expression imaging, an exogenous reporter gene can be incorporated into the genome of a tumor cell line, and its expression can be controlled by a promoter. Eventually, the expression products of the reporter gene by the promoter can be used as contrast for imaging, either directly or indirectly via some assay. Because many diseases like cancer are related to genetic disorders, imaging of gene

Table 2. Comparison of the Optical Cross Sections of Au-Based Nanostructures Obtained Experimentally by Photoacoustic Imaging and Theoretical Calculation, Respectively^a

		cross-section (10^{-15} m ²)		
		extinction (σ_e)	absorption (σ_a)	σ_a/σ_e
nanocages (45.0 nm)	experimental	7.26 ± 0.06	5.96 ± 0.25	0.82 ± 0.04
	calculated	19.88	16.26	0.82
nanocages (32.0 nm)	experimental	2.99 ± 0.04	3.05 ± 0.12	1 ± 0.04
	calculated	6.39	6.02	0.94
nanorods (44.0×19.8 nm ²)	experimental	2.16 ± 0.02	1.87 ± 0.23	0.87 ± 0.11
	calculated	2.15	1.83	0.85
nanospheres (150 nm)	experimental	145 ± 14.76		
	calculated	99.96	5.73	0.056

^a Reprinted with permission from ref 126. Copyright 2009 American Chemical Society.




Figure 50. Optical spectra of the blue product, oxyhemoglobin (HbO_2), and deoxyhemoglobin (Hb). Reprinted with permission from ref 127. Copyright 2007 Society of Photo-Optical Instrumentation Engineers.

expression could potentially play an important role in molecular imaging.

The dark-field confocal PAM (50 MHz, see section 2.2.1) was employed to image the expression of lacZ reporter gene in a brain of a Sprague–Dawley rat (80–100 g) *in vivo*. The lacZ reporter gene encodes β -galactosidase, an *E. coli* enzyme for lactose metabolism. 5-Bromo-4-chloro-3-indolyl- β -D-galactoside (X-gal), a sensitive colorimetric assay, was used for β -galactosidase staining. A blue product is created when β -galactosidase cleaves the X-gal's glycosidic linkage, and this blue product has strong optical absorption (Figure 50). Consequently, this provides strong contrast for PA imaging.

9 L/lacZ gliosarcoma tumor cells were inoculated to the scalp of a rat. When the tumor was visible, 20 μL of an X-gal solution (20 mg/mL) was administered near the tumor 1 day before PA imaging. Figure 51A shows the PA imaging at an optical wavelength of 635 nm of the lacZ-marked 9 L gliosarcoma after X-gal staining. The tumor, enhanced by blue product, is clearly seen. Without X-gal staining, the tumor was not observable in the PA image. The surrounding microvasculature was imaged at the wavelength of 584 nm in Figure 51B. Figure 51C shows a composite image of both blood vessels (Figure 51B) and the tumor (Figure 51A) with enhanced contrast by the blue product.

4.4. Fluorescence Proteins¹²⁹

Fluorescence protein imaging has become increasingly important in biological and medical research.¹³⁰ However, because of the shallow imaging depth (~ 1 mm) of fluorescence microscopy, the applicability of fluorescence proteins is quite limited. As mentioned in section 4.1, PA signals are proportional to the product of the molar extinction coefficient and the nonradiative quantum yield ($1 - \text{fluorescence}$

Table 3. Optical Properties of Fluorescence Proteins^{131a}

FPs	EX [nm]	EM [nm]	molar extinction coefficient [$\text{cm}^{-1} \text{M}^{-1}$]	FQY [%]	molar extinction coefficient \times NR-QY [$\text{cm}^{-1} \text{M}^{-1}$]
mCherry	587	610	0.72×10^5	22	0.56×10^5
tdTomato	554	581	1.38×10^5	69	0.42×10^5
mStrawberry	574	596	0.9×10^5	29	0.64×10^5
J-Red	584	610	0.44×10^5	20	0.35×10^5
DsRed-Monomer	556	586	0.35×10^5	10	0.32×10^5
EGFP	488	507	0.56×10^5	60	0.22×10^5

^a FPs, fluorescence proteins; EX, excitation wavelength; EM, emission wavelength; FQY, fluorescence quantum yield; and NR-QY, nonradiative quantum yield ($1 - \text{FQY}$).

quantum yield) of a contrast agent. Thus, fluorescence proteins possessing less fluorescent quantum yields can be great candidates as contrast agents for PA imaging. Furthermore, because a deeper imaging depth is possible with PA imaging, molecular PA imaging with fluorescence proteins has great potential in biology and medicine. Table 3 summarizes the optical properties of several fluorescence proteins.

A single-element cylindrically focused ultrasound transducer-based PACT system¹²⁹ was employed to image the expression of mCherry in the head of an adult transgenic zebrafish *in vivo*. Figure 52A shows the morphological cross-sectional PA images of the head at various depths. The structural PA image in Figure 52B matches well with the corresponding histology in Figure 52C. The map of mCherry expression was overlaid on the structural PA image taken at 585 nm in Figure 52D. The mCherry expression is clearly seen in the PA image, which well correlates with the corresponding epifluorescence image of the dissected brain (Figure 52E).

5. Summary

In summary, two types of PAT modalities have been reviewed: (1) inverse-algorithm-based PACT and (2) direct-raster scanning-based PAM. PAT has the following features: (1) PAT breaks through the fundamental imaging depth limit of high-resolution optical microscopy. Both the imaging depth and the spatial resolution are scalable with ultrasonic frequency and can be tuned for specific imaging requirements. (2) The image contrast of PAT is primarily derived from optical absorption, unlike pure optical microscopy. (3) With ultrasonic arrays, PAT can acquire images in real time. (4) PAT uses nonionizing radiation, is safe for humans, and is ready for use in clinical applications. (5) PAT provides morphological and functional (oxygen saturation of hemoglobin and total hemoglobin concentration) imaging using intrinsic contrast (hemoglobin or melanin). (6) PAT provides functional (sentinel lymph node mapping) and molecular




Figure 51. *In vivo* noninvasive PA images of lacZ-marked tumor. (A) PA image acquired at 635 nm showing the tumor morphology after X-gal staining. (B) PA image acquired at 584 nm showing microvasculature. (C) A composite image of the surrounding vasculature (B) and the primary tumor (A). Reprinted with permission from ref 128. Copyright 2008 World Scientific.




Figure 52. Three-dimensional in vivo PA images of the brain of an adult mCherry-expressing transgenic zebrafish. (A) Five transverse PA images taken at 585 nm. (B) A PA image and (C) its corresponding histological section. DM, dorsal fin musculature; B, hindbrain; N, lateral line nerve; O, operculum; S, skull bones; P, pharynx; H, heart; HM, hypobranchial musculature. (D) PA image of the brain with mCherry expression. (E) Corresponding fluorescent histology of a dissected brain. Reprinted with permission from ref 129. Copyright 2009 Nature.

(tumor specific) imaging using exogenous contrast agents (organic dyes, nanoparticles, reporter genes, or fluorescent proteins).

6. Acknowledgments

We appreciate Professor James Ballard's close reading of the manuscript and Song Hu for cover art. This work was sponsored in part by National Institutes of Health grants R01 EB000712, R01 EB008085, R01 CA113453901, U54 CA136398, and 5P60 DK02057933. L.V.W. has a financial interest in Microphotoacoustics, Inc., and Endra, Inc., which, however, did not support this work.

7. References

- Bell, A. G. *Am. J. Sci.* **1880**, *20*, 305.
- Denk, W.; Strickler, J. H.; Webb, W. W. *Science* **1990**, *248*, 73.
- Huang, D.; Swanson, E. A.; Lin, C. P.; Schuman, J. S.; Stinson, W. G.; Chang, W.; Hee, M. R.; Flotte, T.; Gregory, K.; Puliafito, C. P.; Fujimoto, J. G. *Science* **1991**, *254*, 1178.
- Singh, A.; Gopinathan, K. P. *Curr. Sci. India* **1998**, *74*, 841.
- Zeff, B. W.; White, B. R.; Dehghani, H.; Schlaggar, B. L.; Culver, J. P. *Proc. Natl. Acad. Sci. U.S.A.* **2007**, *104*, 12169.
- Wang, X. D.; Pang, Y. J.; Ku, G.; Xie, X. Y.; Stoica, G.; Wang, L. H. V. *Nat. Biotechnol.* **2003**, *21*, 803.
- Gamelin, J.; Maurudis, A.; Aguirre, A.; Huang, F.; Guo, P.; Wang, L. V.; Zhu, Q. *Opt. Express* **2009**, *17*, 10489.
- Gamelin, J.; Aguirre, A.; Maurudis, A.; Huang, F.; Castillo, D.; Wang, L. V.; Zhu, Q. *J. Biomed. Opt.* **2008**, *13*, 024007.
- Zhang, H. F.; Maslov, K.; Stoica, G.; Wang, L. H. V. *Nat. Biotechnol.* **2006**, *24*, 848.
- Maslov, K.; Stoica, G.; Wang, L. V. *Opt. Lett.* **2005**, *30*, 625.
- Song, K. H.; Wang, L. V. *J. Biomed. Opt.* **2007**, *12*, 060503.
- Maslov, K.; Zhang, H. F.; Hu, S.; Wang, L. V. *Opt. Lett.* **2008**, *33*, 929.
- Ku, G.; Wang, X. D.; Xie, X. Y.; Stoica, G.; Wang, L. H. V. *Appl. Opt.* **2005**, *44*, 770.
- Song, K. H.; Wang, L. V. *Med. Phys.* **2008**, *35*, 4524.
- Wang, X.; Xie, X.; Ku, G.; Wang, L. V.; Stoica, G. *J. Biomed. Opt.* **2006**, *11*, 024015.
- Zhang, H. F.; Maslov, K.; Sivaramakrishnan, M.; Stoica, G.; Wang, L. H. V. *Appl. Phys. Lett.* **2007**, *90*, 053901.
- Zemp, R. J.; Song, L.; Bitton, R.; Shung, K. K.; Wang, L. V. *Opt. Express* **2008**, *16*, 18551.
- Li, P. C.; Wang, C. R. C.; Shieh, D. B.; Wei, C. W.; Liao, C. K.; Poe, C.; Jhan, S.; Ding, A. A.; Wu, Y. N. *Opt. Express* **2008**, *16*, 18605.
- De la Zerna, A.; Zavaleta, C.; Keren, S.; Vaithilingam, S.; Bodapati, S.; Liu, Z.; Levi, J.; Smith, B. R.; Ma, T. J.; Oralkan, O.; Cheng, Z.; Chen, X.; Dai, H.; Khuri-Yakub, B. T.; Gambhir, S. S. *Nat. Nanotechnol.* **2008**, *3*, 557.
- Xu, M.; Wang, L. V. *Phys. Rev. E* **2005**, *71*, 016706.
- Kostli, K. P.; Frenz, M.; Weber, H. P.; Paltauf, G.; Schmidt-Kloiber, H. *Appl. Opt.* **2001**, *40*, 3800.
- American National Standard for Safe Use of Lasers (ANSI Z136.1-2000)*; Laser Institute of America: 2000.
- Song, L.; K, M.; Bitton, R.; Shung, K. K.; Wang, L. V. *J. Biomed. Opt.* **2008**, *13*, 054028.
- Bitton, R.; Zemp, R. J.; Yen, J.; Wang, L. V.; Shung, K. K. *IEEE T. Med. Imaging* **2009**, *28*, 1190.
- Zhang, H. F.; Maslov, K.; Wang, L. V. *Nat. Protoc.* **2007**, *2*, 797.
- Oravsky, A. A.; Karabutov, A. A. In *Biomedical Photonics Handbook*; Vo-Dinh, T., Ed.; CRC Press: Boca Raton, FL, 2003; Vol. PMT125.
- Stein, E. W.; Maslov, K.; Wang, L. V. *J. Appl. Phys.* **2009**, *105*, 102027.
- Zharov, V. P.; Galanzha, E. I.; Shashkov, E. V.; Khlebtsov, N. G.; Tuchin, V. V. *Opt. Lett.* **2007**, *31*, 3623.
- Wang, L. V.; Wu, H. I. *Biomedical Optics: Principles and Imaging*; John Wiley & Sons, Inc.: Hoboken, NJ, 2007.
- Vanzetta, I.; Grinvald, A. *Science* **1999**, *286*, 1555.
- Chandrakala, C.; Fraker, D. L. *Cancer Lett.* **2005**, *211*, 225.
- Tandara, A. A.; Mustoe, T. A. *World J. Surg.* **2004**, *28*, 294.
- Foo, S. S.; Abbott, D. F.; Lawrentschuk, N.; Scott, A. M. *Mol. Imaging Biol.* **2004**, *6*, 291.
- Chance, B.; Borer, E.; Evans, A.; Holtom, G.; Kent, J.; Maris, M.; McCully, K.; Northrop, J.; Shinkwin, M. *Ann. N.Y. Acad. Sci.* **1988**, *551*, 1.
- Prasad, P. V.; Edelman, R. R.; Epstein, F. H. *Circulation* **1996**, *94*, 3271.
- Kuppusamy, P.; Afeworki, M.; Shankar, R. A.; Coffin, D.; Krishna, M. C.; Hahn, S. M.; Mitchell, J. B.; Zweier, J. L. *Cancer Res.* **1998**, *58*, 1562.
- Cher, L. M.; Murone, C.; Lawrentschuk, N.; Ramdave, S.; Papenfuss, A.; Hannah, A.; O'Keefe, G. J.; Sachinidis, J. I.; Berlangieri, S. U.; Fabinyi, G.; Scott, A. M. *J. Nucl. Med.* **2006**, *47*, 410.
- Li, L.; Yu, J. M.; Xing, L. G.; Ma, K. F.; Zhu, H.; Guo, H. B.; Sun, X. D.; Li, J. B.; Yang, G. R.; Li, W. L.; Yue, J. B.; Li, B. S. *Am. J. Clin. Oncol.-Cancer* **2006**, *29*, 628.
- Jensen, D. *The Principles of Physiology*; Appleton-Century-Crofts: New York, 1976.
- Duong, T. Q.; Iadecola, C.; Kim, S. G. *Magn. Reson. Med.* **2001**, *45*, 61.
- Stein, E. W.; Maslov, K.; Wang, L. V. *H. J. Biomed. Opt.* **2009**, *14*, 020520.
- Li, M. L.; Oh, J. T.; Xie, X. Y.; Ku, G.; Wang, W.; Li, C.; Lungu, G.; Stoica, G.; Wang, L. V. *Proc. IEEE* **2008**, *96*, 481.
- Hanahan, D.; Weinberg, R. A. *Cell* **2000**, *100*, 57.
- Bergers, G.; Benjamin, L. E. *Nat. Rev. Cancer* **2003**, *3*, 401.
- Hall, D. G.; Stoica, G. *Clin. Exp. Metastasis* **1994**, *12*, 283.
- Maeda, H.; Fang, J.; Inutsuka, T.; Kitamoto, Y. *Int. Immunopharmacol.* **2003**, *3*, 319.
- Vaupel, P.; Harrison, L. *Oncologist* **2004**, *9*, 4.
- Vaupel, P.; Thews, O.; Hoekel, M. *Med. Oncol.* **2001**, *18*, 243.
- Lal, S.; Lacroix, M.; Tofilon, P.; Fuller, G. N.; Sawaya, R.; Lang, F. F. *J. Neurosurg.* **2000**, *92*, 326.
- Hu, S.; Maslov, K.; Wang, L. H. V. *Med. Phys.* **2009**, *36*, 2320.
- Jain, R. K.; Munn, L. L.; Fukumura, D. *Nat. Rev. Cancer* **2002**, *2*, 266.
- Fukumura, D.; Jain, R. K. *Apmis* **2008**, *116*, 695.
- Deisseroth, K.; Feng, G. P.; Majewska, A. K.; Miesenbock, G.; Ting, A.; Schnitzer, M. J. *J. Neurosci.* **2006**, *26*, 10380.
- Zepeda, A.; Arias, C.; Sengpiel, F. *J. Neurosci. Methods* **2004**, *136*, 1.
- Chin, G. A.; Diegelmann, R. F.; Schultx, G. S. *Wound Healing*; Taylor & Francis: Boca Raton, FL, 2005.
- Ku, G.; Wang, L. V. *Opt. Lett.* **2005**, *30*, 507.
- Wang, X.; Ku, G.; Wegiel, M. A.; Bornhop, D. J.; Stoica, G.; Wang, L. V. *Opt. Lett.* **2004**, *29*, 730.
- Luetkemeier, M. J.; Fattor, J. A. *Clin. Chem.* **2001**, *47*, 1843.
- Reindl, S.; Penzkofer, A.; Gong, S. H.; Landthaler, M.; Szeimies, R. M.; Abels, C.; Bäuml, W. *J. Photochem. Photobiol., A* **1997**, *105*, 65.
- Landsman, M. L.; Kwant, G.; Mook, G. A.; Zijlstra, W. G. *J. Appl. Physiol.* **1976**, *40*, 575.
- Zalipsky, S.; Brandeis, E.; Newman, M. S.; Woodle, M. C. *FEBS Lett.* **1994**, *353*, 71.

- (62) LiCor.com; <http://www.licor.com> (accessed July 3, 2009), IRDye Infrared Dyes: Nebraska.
- (63) Wang, W.; Ke, S.; Wu, Q.; Charnsangavej, C.; Gurfinkel, M.; Gelovani, J. G.; Abbruzzese, J. L.; Sevick-Muraca, E. M.; Li, C. *Mol. Imaging* **2004**, *3*, 343.
- (64) Brooks, P. C.; Clark, R. A.; Cheresch, D. A. *Science* **1994**, *264*, 569.
- (65) Song, K. H.; Stein, E. W.; Margenthaler, J. A.; Wang, L. V. *J. Biomed. Opt.* **2008**, *13*, 054033.
- (66) Prah, S. <http://omlc.ogi.edu/> (accessed July 03, 2009), Methylene blue, 2007.
- (67) Wikipedia.org; <http://en.wikipedia.org> (accessed July 03, 2009), Methylene blue.
- (68) Abe, H.; Wagner, S. J.; Kuwabara, M.; Kamo, N.; Ikebuchi, K.; Sekiguchi, S. *Photochem. Photobiol.* **1997**, *65*, 873.
- (69) Schneider, J. E., Jr.; Tabatabaie, T.; Maidt, L.; Smith, R. H.; Nguyen, X.; Pye, Q.; Floyd, R. A. *Photochem. Photobiol.* **1998**, *67*, 350.
- (70) Kobayashi, H.; Kawamoto, S.; Sakai, Y.; Choyke, P. L.; Star, R. A.; Brechbiel, M. W.; Sato, N.; Tagaya, Y.; Morris, J. C.; Waldmann, T. A. *J. Natl. Cancer Inst.* **2004**, *96*, 703.
- (71) McMasters, K. M.; Tuttle, T. M.; Carlson, D. J.; Brown, C. M.; Noyes, R. D.; Glaser, R. L.; Vennekotter, D. J.; Turk, P. S.; Tate, P. S.; Sardi, A.; Cerrito, P. B.; Edwards, M. J. *J. Clin. Oncol.* **2000**, *18*, 2560.
- (72) Purushotham, A. D.; Upponi, S.; Klevesath, M. B.; Bobrow, L.; Millar, K.; Myles, J. P.; Duffy, S. W. *J. Clin. Oncol.* **2005**, *23*, 4312.
- (73) Krishnamurthy, S.; Sneige, N.; Bedi, D. G.; Edeiken, B. S.; Fornage, B. D.; Kuerer, H. M.; Singletary, S. E.; Hunt, K. K. *Cancer* **2002**, *95*, 982.
- (74) Liu, Y.; Truini, C.; Ariyan, S. *Ann. Surg. Oncol.* **2008**, *15*, 2412.
- (75) Panyam, J.; Labhasetwar, V. *Adv. Drug Delivery Rev.* **2003**, *55*, 329.
- (76) Sahoo, S. K.; Labhasetwar, V. *Drug Discovery Today* **2003**, *8*, 1112.
- (77) Gao, X.; Cui, Y.; Levenson, R. M.; Chung, L. W.; Nie, S. *Nat. Biotechnol.* **2004**, *22*, 969.
- (78) True, L. D.; Gao, X. *J. Mol. Diagn.* **2007**, *9*, 7.
- (79) Harisinghani, M. G.; Barentsz, J.; Hahn, P. F.; Deserno, W. M.; Tabatabaie, S.; van de Kaa, C. H.; de la Rosette, J.; Weissleder, R. *N. Engl. J. Med.* **2003**, *348*, 2491.
- (80) Winter, P. M.; Caruthers, S. D.; Kassner, A.; Harris, T. D.; Chinen, L. K.; Allen, J. S.; Lacy, E. K.; Zhang, H.; Robertson, J. D.; Wickline, S. A.; Lanza, G. M. *Cancer Res.* **2003**, *63*, 5838.
- (81) Kim, S.; Lim, Y. T.; Soltesz, E. G.; De Grand, A. M.; Lee, J.; Nakayama, A.; Parker, J. A.; Mihajljevic, T.; Laurence, R. G.; Dor, D. M.; Cohn, L. H.; Bawendi, M. G.; Frangioni, J. V. *Nat. Biotechnol.* **2004**, *22*, 93.
- (82) Wang, Y. W.; Xie, X. Y.; Wang, X. D.; Ku, G.; Gill, K. L.; O'Neal, D. P.; Stoica, G.; Wang, L. V. *Nano Lett.* **2004**, *4*, 1689.
- (83) Harisinghani, M. *Lancet Oncol.* **2008**, *9*, 814.
- (84) Heesakkers, R. A.; Hovels, A. M.; Jager, G. J.; van den Bosch, H. C.; Witjes, J. A.; Raat, H. P.; Severens, J. L.; Adang, E. M.; van der Kaa, C. H.; Futterer, J. J.; Barentsz, J. *Lancet Oncol.* **2008**, *9*, 850.
- (85) Kobayashi, H.; Brechbiel, M. W. *Curr. Pharm. Biotechnol.* **2004**, *5*, 539.
- (86) Rabin, O.; Manuel Perez, J.; Grimm, J.; Wojtkiewicz, G.; Weissleder, R. *Nat. Mater.* **2006**, *5*, 118.
- (87) Hovels, A. M.; Heesakkers, R. A.; Adang, E. M.; Jager, G. J.; Barentsz, J. O. *Eur. Radiol.* **2004**, *14*, 1707.
- (88) Hu, M.; Chen, J. Y.; Li, Z. Y.; Au, L.; Hartland, G. V.; Li, X. D.; Marquez, M.; Xia, Y. N. *Chem. Soc. Rev.* **2006**, *35*, 1084.
- (89) Song, K. H.; Kim, C.; Copley, C. M.; Xia, Y.; Wang, L. V. *Nano Lett.* **2009**, *9*, 183.
- (90) Yang, X.; Skrabalak, S. E.; Li, Z. Y.; Xia, Y.; Wang, L. V. *Nano Lett.* **2007**, *7*, 3798.
- (91) Chen, J.; Saeki, F.; Wiley, B. J.; Cang, H.; Cobb, M. J.; Li, Z. Y.; Au, L.; Zhang, H.; Kimmey, M. B.; Li, X.; Xia, Y. *Nano Lett.* **2005**, *5*, 473.
- (92) Skrabalak, S. E.; Chen, J.; Au, L.; Lu, X.; Li, X.; Xia, Y. *Adv. Mater.* **2007**, *19*, 3177.
- (93) Cuenca, A. G.; Jiang, H.; Hochwald, S. N.; Delano, M.; Cance, W. G.; Grobmyer, S. R. *Cancer* **2006**, *107*, 459.
- (94) Sun, Y. G.; Xia, Y. N. *J. Am. Chem. Soc.* **2004**, *126*, 3892.
- (95) Chen, J. Y.; McLellan, J. M.; Siekkinen, A.; Xiong, Y. J.; Li, Z. Y.; Xia, Y. N. *J. Am. Chem. Soc.* **2006**, *128*, 14776.
- (96) Skrabalak, S. E.; Chen, J. Y.; Sun, Y. G.; Lu, X. M.; Au, L.; Copley, C. M.; Xia, Y. N. *Acc. Chem. Res.* **2008**, *41*, 1587.
- (97) Sun, Y.; Xia, Y. *Science* **2002**, *298*, 2176.
- (98) Sun, Y. G.; Mayers, B.; Xia, Y. N. *Adv. Mater.* **2003**, *15*, 641.
- (99) Sun, Y. G.; Mayers, B. T.; Xia, Y. N. *Nano Lett.* **2002**, *2*, 481.
- (100) Sun, Y. G.; Xia, Y. A. *Nano Lett.* **2003**, *3*, 1569.
- (101) Siekkinen, A. R.; McLellan, J. M.; Chen, J. Y.; Xia, Y. N. *Chem. Phys. Lett.* **2006**, *432*, 491.
- (102) Skrabalak, S. E.; Au, L.; Li, X. D.; Xia, Y. *Nat. Protoc.* **2007**, *2*, 2182.
- (103) Wiley, B.; Herricks, T.; Sun, Y. G.; Xia, Y. N. *Nano Lett.* **2004**, *4*, 2057.
- (104) Chen, J. Y.; Wiley, B.; Li, Z. Y.; Campbell, D.; Saeki, F.; Cang, H.; Au, L.; Lee, J.; Li, X. D.; Xia, Y. N. *Adv. Mater.* **2005**, *17*, 2255.
- (105) Li, M. L.; Wang, J. C.; Schwartz, J. A.; Gill-Sharp, K. L.; Stoica, G.; Wang, L. H. V. *J. Biomed. Opt.* **2009**, *14*, 010507.
- (106) Mallidi, S.; Larson, T.; Aaron, J.; Sokolov, K.; Emelianov, S. *Opt. Express* **2007**, *15*, 6583.
- (107) Oldenburg, S. J.; Jackson, J. B.; Westcott, S. L.; Halas, N. J. *Appl. Phys. Lett.* **1999**, *75*, 2897.
- (108) O'Neal, D. P.; Hirsch, L. R.; Halas, N. J.; Payne, J. D.; West, J. L. *Cancer Lett.* **2004**, *209*, 171.
- (109) Aaron, J.; Nitin, N.; Travis, K.; Kumar, S.; Collier, T.; Park, S. Y.; Jose-Yacaman, M.; Coghlan, L.; Follen, M.; Richards-Kortum, R.; Sokolov, K. *J. Biomed. Opt.* **2007**, *12*, 034007.
- (110) Agarwal, A.; Huang, S. W.; O'Donnell, M.; Day, K. C.; Day, M.; Kotov, N.; Ashkenazi, S. *J. Appl. Phys.* **2007**, *102*, 064701.
- (111) Eghtedari, M.; Oraevsky, A.; Copland, J. A.; Kotov, N. A.; Conjusteau, A.; Motamedi, M. *Nano Lett.* **2007**, *7*, 1914.
- (112) Kim, K.; Huang, S. W.; Ashkenazi, S.; O'Donnell, M.; Agarwal, A.; Kotov, N. A.; Denny, M. F.; Kaplan, M. J. *Appl. Phys. Lett.* **2007**, *90*, 223901.
- (113) Song, K. H.; Kim, C.; Maslov, K.; Wang, L. V. *Eur. J. Radiol.* **2009**, *70*, 227.
- (114) Huang, X. H.; El-Sayed, I. H.; Qian, W.; El-Sayed, M. A. *J. Am. Chem. Soc.* **2006**, *128*, 2115.
- (115) Klimiuk, P. A.; Sierakowski, S.; Latosiewicz, R.; Cylwik, J. P.; Cylwik, B.; Skowronski, J.; Chwiecko, J. *Ann. Rheum. Dis.* **2002**, *61*, 804.
- (116) Pramanik, M.; Song, K. H.; Swierczewska, M.; Green, D.; Sitharaman, B.; Wang, L. V. *Phys. Med. Biol.* **2009**, *54*, 3291.
- (117) Pramanik, M.; Swierczewska, M.; Green, D.; Sitharaman, B.; Wang, L. V. *J. Biomed. Opt.* **2009**, *14*, 034018.
- (118) Xiang, L.; Yuan, Y.; Xing, D.; Ou, Z.; Yang, S.; Zhou, F. *J. Biomed. Opt.* **2009**, *14*, 021008.
- (119) Schipper, M. L.; Nakayama-Ratchford, N.; Davis, C. R.; Kam, N. W. S.; Chu, P.; Liu, Z.; Sun, X. M.; Dai, H. J.; Gambhir, S. S. *Nat. Nanotechnol.* **2008**, *3*, 216.
- (120) Wahrheit, D. B.; Laurence, B. R.; Reed, K. L.; Roach, D. H.; Reynolds, G. A.; Webb, T. R. *Toxicol. Sci.* **2004**, *77*, 117.
- (121) Zhao, Y. L.; Xing, G. M.; Chai, Z. F. *Nat. Nanotechnol.* **2008**, *3*, 191.
- (122) Zhou, F.; Xing, D.; Ou, Z.; Wu, B.; Resasco, D. E.; Chen, W. R. *J. Biomed. Opt.* **2009**, *14*, 021009.
- (123) Liu, Z.; Cai, W. B.; He, L. N.; Nakayama, N.; Chen, K.; Sun, X. M.; Chen, X. Y.; Dai, H. J. *Nat. Nanotechnol.* **2007**, *2*, 47.
- (124) Pan, D. P.; Pramanik, M.; Senpan, A.; Yang, X. M.; Song, K. H.; Scott, M. J.; Zhang, H. Y.; Gaffney, P. J.; Wickline, S. A.; Wang, L. V.; Lanza, G. M. *Angew. Chem., Int. Ed.* **2009**, *48*, 4170.
- (125) Bouchard, L. S.; Anwar, M. S.; Liu, G. L.; Hann, B.; Xie, Z. H.; Gray, J. W.; Wang, X.; Pines, A.; Chen, F. F. *Proc. Natl. Acad. Sci. U.S.A.* **2009**, *106*, 4085.
- (126) Cho, E. C.; Kim, C.; Zhou, F.; Copley, C. M.; Song, K. H.; Chen, J. Y.; Li, Z. Y.; Wang, L. H. V.; Xia, Y. N. *J. Phys. Chem. C* **2009**, *113*, 9023.
- (127) Li, L.; Zemp, R. J.; Lungu, G.; Stoica, G.; Wang, L. V. *J. Biomed. Opt.* **2007**, *12*, 020504.
- (128) Li, L.; Zhang, H. F.; Zemp, R. J.; Maslov, K.; Wang, L. V. *J. Innovative Opt. Health Sci.* **2008**, *1*, 207.
- (129) Razansky, D.; Distel, M.; Vinegoni, C.; Ma, R.; Perrimon, N.; Koster, R. W.; Ntziachristos, V. *Nat. Photonics* **2009**, *3*, 412.
- (130) Giepmans, B. N. G.; Adams, S. R.; Ellisman, M. H.; Tsien, R. Y. *Science* **2006**, *312*, 217.
- (131) Shaner, N. C.; Steinbach, P. A.; Tsien, R. Y. *Nat. Methods* **2005**, *2*, 905.